

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Lockwood, Kessler & Bartlett, Inc. **level:** low **sdg no.:** P4548

contract: LOCK01 **lab code:** CHEM **case no.:** P4548 **sas no.:** P4548

matrix: Water **sample id:** P4548-01 **client id:** MW-1MS

Percent Solids for Sample: NA **Spiked ID:** P4548-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	355		25.0	U	400	89		P
Arsenic	ug/L	75 - 125	391		10.0	U	400	98		P
Barium	ug/L	75 - 125	158		62.5		100	95		P
Beryllium	ug/L	75 - 125	80.7		3.00	U	100	81		P
Cadmium	ug/L	75 - 125	92.7		2.09	J	100	91		P
Chromium	ug/L	75 - 125	177		1.48	J	200	88		P
Cobalt	ug/L	75 - 125	93.1		15.0	U	100	93		P
Copper	ug/L	75 - 125	138		10.0	U	150	92		P
Iron	ug/L	75 - 125	1590		42.8	J	1500	103		P
Lead	ug/L	75 - 125	388		6.00	U	500	78		P
Manganese	ug/L	75 - 125	179		86.6		100	93		P
Nickel	ug/L	75 - 125	243		6.80	J	250	95		P
Potassium	ug/L	75 - 125	10200		5230		5000	99		P
Selenium	ug/L	75 - 125	804		10.0	U	1000	80		P
Silver	ug/L	75 - 125	22.5		5.00	U	37.5	60	N	P
Sodium	ug/L	75 - 125	54400		53100		1500	87		P
Thallium	ug/L	75 - 125	862		20.0	U	1000	86		P
Vanadium	ug/L	75 - 125	150		20.0	U	150	100		P
Zinc	ug/L	75 - 125	298		254		100	43	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Lockwood, Kessler & Bartlett, Inc. **level:** low **sdg no.:** P4548

contract: LOCK01 **lab code:** CHEM **case no.:** P4548 **sas no.:** P4548

matrix: Water **sample id:** P4548-01 **client id:** MW-1MSD

Percent Solids for Sample: NA **Spiked ID:** P4548-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	367		25.0	U	400	92		P
Arsenic	ug/L	75 - 125	401		10.0	U	400	100		P
Barium	ug/L	75 - 125	161		62.5		100	99		P
Beryllium	ug/L	75 - 125	83.1		3.00	U	100	83		P
Cadmium	ug/L	75 - 125	95.0		2.09	J	100	93		P
Chromium	ug/L	75 - 125	181		1.48	J	200	90		P
Cobalt	ug/L	75 - 125	95.5		15.0	U	100	96		P
Copper	ug/L	75 - 125	141		10.0	U	150	94		P
Iron	ug/L	75 - 125	1580		42.8	J	1500	103		P
Lead	ug/L	75 - 125	398		6.00	U	500	80		P
Manganese	ug/L	75 - 125	184		86.6		100	98		P
Nickel	ug/L	75 - 125	249		6.80	J	250	97		P
Potassium	ug/L	75 - 125	10200		5230		5000	99		P
Selenium	ug/L	75 - 125	826		10.0	U	1000	83		P
Silver	ug/L	75 - 125	22.8		5.00	U	37.5	61	N	P
Sodium	ug/L	75 - 125	54400		53100		1500	87		P
Thallium	ug/L	75 - 125	831		20.0	U	1000	83		P
Vanadium	ug/L	75 - 125	153		20.0	U	150	102		P
Zinc	ug/L	75 - 125	303		254		100	49	N	P