

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** P4548
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** P4548 **SAS No.:** P4548

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164438BS							
Antimony	ug/L	400	375		94	80 - 120	P
Arsenic	ug/L	400	397		99	80 - 120	P
Barium	ug/L	100	99.2		99	80 - 120	P
Beryllium	ug/L	100	92.9		93	80 - 120	P
Cadmium	ug/L	100	84.8		85	80 - 120	P
Chromium	ug/L	200	188		94	80 - 120	P
Cobalt	ug/L	100	89.4		89	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1580		105	80 - 120	P
Lead	ug/L	500	426		85	80 - 120	P
Manganese	ug/L	100	101		101	80 - 120	P
Nickel	ug/L	250	225		90	80 - 120	P
Potassium	ug/L	5000	5090		102	80 - 120	P
Selenium	ug/L	1000	979		98	80 - 120	P
Silver	ug/L	37.5	33.9		90	80 - 120	P
Sodium	ug/L	1500	1380		92	80 - 120	P
Thallium	ug/L	1000	903		90	80 - 120	P
Vanadium	ug/L	150	157		105	80 - 120	P
Zinc	ug/L	100	89.1		89	80 - 120	P