

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: LOCK01

Lab Code: CHEM Case No.: P4548 SAS No.: P4548 SDG No.: P4548

Instrument ID: MSVOA\_X Calibration Date(s): 10/28/2024 10/28/2024

Heated Purge: (Y/N) N Calibration Time(s): 10:59 12:54

GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:              |        | RRF001 = VX043556.D |        | RRF005 = VX043557.D |        | RRF020 = VX043558.D |       |       |  |
|---------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                           |        | RRF050 = VX043559.D |        | RRF100 = VX043560.D |        | RRF150 = VX043561.D |       |       |  |
| COMPOUND                  | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| Vinyl Chloride            | 0.626  | 0.674               | 0.638  | 0.634               | 0.643  | 0.591               | 0.634 | 4.2   |  |
| Chloroethane              | 0.263  | 0.215               | 0.212  | 0.225               | 0.217  | 0.210               | 0.223 | 9     |  |
| Trichlorofluoromethane    | 0.823  | 0.776               | 0.710  | 0.779               | 0.846  | 0.716               | 0.775 | 7.1   |  |
| 1,1-Dichloroethene        | 0.533  | 0.555               | 0.521  | 0.536               | 0.556  | 0.513               | 0.536 | 3.3   |  |
| Acrylonitrile             | 0.302  | 0.335               | 0.331  | 0.348               | 0.360  | 0.340               | 0.336 | 5.8   |  |
| Acetone                   | 0.365  | 0.320               | 0.294  | 0.312               | 0.314  | 0.295               | 0.317 | 8.2   |  |
| Carbon Disulfide          | 1.067  | 0.974               | 1.007  | 1.134               | 1.310  | 1.227               | 1.120 | 11.6  |  |
| Methylene Chloride        | 0.772  | 0.699               | 0.663  | 0.663               | 0.665  | 0.633               | 0.682 | 7.2   |  |
| Vinyl Acetate             | 1.326  | 1.547               | 1.607  | 1.688               | 1.706  | 1.618               | 1.582 | 8.7   |  |
| 2-Butanone                | 0.427  | 0.462               | 0.457  | 0.483               | 0.484  | 0.459               | 0.462 | 4.5   |  |
| Carbon Tetrachloride      | 0.462  | 0.442               | 0.423  | 0.444               | 0.465  | 0.429               | 0.444 | 3.8   |  |
| cis-1,2-Dichloroethene    | 0.708  | 0.719               | 0.716  | 0.740               | 0.747  | 0.707               | 0.723 | 2.3   |  |
| Bromochloromethane        | 0.512  | 0.529               | 0.531  | 0.507               | 0.545  | 0.516               | 0.524 | 2.7   |  |
| Chloroform                | 1.171  | 1.170               | 1.165  | 1.163               | 1.166  | 1.087               | 1.154 | 2.8   |  |
| 1,1,1-Trichloroethane     | 0.939  | 0.949               | 0.978  | 0.986               | 1.020  | 0.939               | 0.969 | 3.3   |  |
| Benzene                   | 1.335  | 1.396               | 1.383  | 1.355               | 1.333  | 1.246               | 1.341 | 4     |  |
| 1,2-Dichloropropane       | 0.327  | 0.339               | 0.325  | 0.338               | 0.333  | 0.320               | 0.330 | 2.2   |  |
| Dibromomethane            | 0.230  | 0.264               | 0.258  | 0.260               | 0.262  | 0.252               | 0.254 | 4.9   |  |
| Bromodichloromethane      | 0.378  | 0.415               | 0.435  | 0.476               | 0.495  | 0.477               | 0.446 | 10    |  |
| 4-Methyl-2-Pentanone      | 0.416  | 0.497               | 0.507  | 0.525               | 0.518  | 0.491               | 0.492 | 8     |  |
| Toluene                   | 0.743  | 0.832               | 0.866  | 0.838               | 0.831  | 0.769               | 0.813 | 5.7   |  |
| t-1,3-Dichloropropene     | 0.421  | 0.424               | 0.485  | 0.509               | 0.526  | 0.510               | 0.479 | 9.6   |  |
| cis-1,3-Dichloropropene   | 0.393  | 0.495               | 0.535  | 0.543               | 0.560  | 0.536               | 0.510 | 12    |  |
| 2-Hexanone                | 0.333  | 0.369               | 0.380  | 0.400               | 0.392  | 0.368               | 0.374 | 6.3   |  |
| Dibromochloromethane      | 0.259  | 0.289               | 0.332  | 0.367               | 0.387  | 0.377               | 0.335 | 15.4  |  |
| 1,2-Dibromoethane         | 0.314  | 0.351               | 0.361  | 0.367               | 0.369  | 0.346               | 0.351 | 5.8   |  |
| Tetrachloroethene         | 0.337  | 0.323               | 0.327  | 0.309               | 0.308  | 0.284               | 0.314 | 5.8   |  |
| Chlorobenzene             | 1.098  | 1.110               | 1.083  | 1.075               | 1.081  | 1.007               | 1.076 | 3.3   |  |
| 1,1,1,2-Tetrachloroethane | 0.301  | 0.345               | 0.350  | 0.372               | 0.381  | 0.359               | 0.351 | 8.1   |  |
| Ethyl Benzene             | 1.757  | 1.763               | 1.799  | 1.777               | 1.794  | 1.634               | 1.754 | 3.5   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: LOCK01  
 Lab Code: CHEM Case No.: P4548 SAS No.: P4548 SDG No.: P4548  
 Instrument ID: MSVOA\_X Calibration Date(s): 10/28/2024 10/28/2024  
 Heated Purge: (Y/N) N Calibration Time(s): 10:59 12:54  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                |        | RRF001 = VX043556.D |        | RRF005 = VX043557.D |        | RRF020 = VX043558.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF050 = VX043559.D |        | RRF100 = VX043560.D |        | RRF150 = VX043561.D |       |       |  |
| COMPOUND                    | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| m/p-Xylenes                 | 0.625  | 0.680               | 0.708  | 0.691               | 0.691  | 0.629               | 0.671 | 5.2   |  |
| o-Xylene                    | 0.616  | 0.698               | 0.698  | 0.689               | 0.698  | 0.645               | 0.674 | 5.2   |  |
| Styrene                     | 1.007  | 1.055               | 1.174  | 1.183               | 1.189  | 1.114               | 1.120 | 6.8   |  |
| Bromoform                   | 0.203  | 0.206               | 0.230  | 0.275               | 0.304  | 0.297               | 0.253 | 17.9  |  |
| 1,1,2,2-Tetrachloroethane   | 1.210  | 1.355               | 1.316  | 1.292               | 1.276  | 1.193               | 1.274 | 4.9   |  |
| 1,2,3-Trichloropropane      | 0.928  | 1.210               | 1.166  | 1.071               | 1.043  | 0.996               | 1.069 | 9.8   |  |
| 1,4-Dichlorobenzene         | 1.917  | 1.720               | 1.694  | 1.653               | 1.657  | 1.551               | 1.699 | 7.1   |  |
| 1,2-Dichlorobenzene         | 1.638  | 1.735               | 1.698  | 1.726               | 1.675  | 1.586               | 1.676 | 3.4   |  |
| 1,2-Dibromo-3-Chloropropane | 0.229  | 0.220               | 0.256  | 0.281               | 0.291  | 0.281               | 0.260 | 11.5  |  |
| 1,2-Dichloroethane-d4       |        | 0.902               | 0.794  | 0.785               | 0.771  | 0.735               | 0.797 | 7.9   |  |
| Dibromofluoromethane        |        | 0.358               | 0.337  | 0.336               | 0.342  | 0.323               | 0.339 | 3.7   |  |
| Toluene-d8                  |        | 1.255               | 1.206  | 1.175               | 1.159  | 1.072               | 1.174 | 5.8   |  |
| 4-Bromofluorobenzene        |        | 0.446               | 0.431  | 0.444               | 0.439  | 0.415               | 0.435 | 2.9   |  |
| Methyl Iodide               |        | 0.726               | 0.762  | 0.782               | 0.789  | 0.726               | 0.757 | 4     |  |
| trans-1,4-Dichloro-2-butene |        | 0.290               | 0.356  | 0.418               | 0.447  | 0.432               | 0.389 | 16.7  |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.