

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043561.D
 Acq On : 28 Oct 2024 12:54
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Oct 28 13:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:29:33 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	264800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	318806	138.969	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 277.940%#			
35) Dibromofluoromethane	5.385	113	256339	145.381	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 290.760%#			
50) Toluene-d8	8.647	98	851731	139.313	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 278.620%#			
62) 4-Bromofluorobenzene	11.079	95	329323	147.475	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 294.940%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	210421	130.533	ug/l	99
3) Chloromethane	1.295	50	265669	129.182	ug/l	100
4) Vinyl Chloride	1.374	62	256442	136.893	ug/l	98
5) Bromomethane	1.593	94	112014	153.092	ug/l	95
6) Chloroethane	1.660	64	90954	148.384	ug/l	100
7) Trichlorofluoromethane	1.861	101	310596	126.333	ug/l	100
8) Diethyl Ether	2.136	74	151359	140.945	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.313	101	212838	137.151	ug/l	99
10) Methyl Iodide	2.441	142	315072	146.614	ug/l	98
11) Tert butyl alcohol	3.026	59	337112	688.018	ug/l #	100
12) 1,1-Dichloroethene	2.307	96	222500	145.831	ug/l	96
13) Acrolein	2.239	56	241097	624.444	ug/l	98
14) Allyl chloride	2.654	41	395090	145.930	ug/l	98
15) Acrylonitrile	3.069	53	736707	678.747	ug/l	99
16) Acetone	2.392	43	639840	653.408	ug/l	99
17) Carbon Disulfide	2.495	76	532448	168.596	ug/l	99
18) Methyl Acetate	2.709	43	392340	133.313	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	867383	146.111	ug/l	97
20) Methylene Chloride	2.782	84	274424	143.401	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	236759	143.334	ug/l	97
22) Diisopropyl ether	3.764	45	790777	141.515	ug/l	96
23) Vinyl Acetate	3.721	43	3509530	760.241	ug/l	99
24) 1,1-Dichloroethane	3.605	63	454844	144.305	ug/l	99
25) 2-Butanone	4.568	43	995055	669.046	ug/l	99
26) 2,2-Dichloropropane	4.471	77	367692	161.856	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	306808	149.076	ug/l	98
28) Bromochloromethane	4.898	49	223906	143.593	ug/l	99
29) Tetrahydrofuran	5.013	42	639748	658.777	ug/l	99
30) Chloroform	5.093	83	471770	139.964	ug/l	99
31) Cyclohexane	5.458	56	333317	133.285	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	407373	143.818	ug/l	99
36) 1,1-Dichloropropene	5.684	75	292944	140.377	ug/l	99
37) Ethyl Acetate	4.721	43	384683	135.067	ug/l	99
38) Carbon Tetrachloride	5.672	117	340536	143.607	ug/l	98
39) Methylcyclohexane	7.379	83	375923	140.469	ug/l	98
40) Benzene	6.031	78	989523	141.848	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043561.D
 Acq On : 28 Oct 2024 12:54
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/29/2024
 Supervised By : Mahesh Dadoda 10/29/2024

Quant Time: Oct 28 13:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:29:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	203044	141.732	ug/l	96
42) 1,2-Dichloroethane	6.086	62	370244	144.284	ug/l	99
43) Isopropyl Acetate	6.342	43	646810	143.566	ug/l	99
44) Trichloroethene	7.123	130	251726	152.200	ug/l	96
45) 1,2-Dichloropropane	7.428	63	254281	149.267	ug/l	99
46) Dibromomethane	7.580	93	200317	153.073	ug/l	98
47) Bromodichloromethane	7.818	83	379264	163.366	ug/l	99
48) Methyl methacrylate	7.696	41	320841	152.537	ug/l	98
49) 1,4-Dioxane	7.671	88	113004	2091.122	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	1948870	682.697	ug/l	99
52) Toluene	8.714	92	611241	142.120	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	405296	172.509	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	426054	165.131	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	259723	143.100	ug/l	99
56) Ethyl methacrylate	9.116	69	449693	153.315	ug/l	97
57) 1,3-Dichloropropane	9.305	76	437811	146.096	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	1091414	756.448	ug/l	98
59) 2-Hexanone	9.433	43	1462736	677.655	ug/l	98
60) Dibromochloromethane	9.519	129	299138	176.144	ug/l	99
61) 1,2-Dibromoethane	9.610	107	275091	150.355	ug/l	99
64) Tetrachloroethene	9.269	164	198596	136.289	ug/l	95
65) Chlorobenzene	10.080	112	703451	145.269	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	250797	153.341	ug/l	100
67) Ethyl Benzene	10.195	91	1141257	141.675	ug/l	100
68) m/p-Xylenes	10.299	106	878867	287.969	ug/l	97
69) o-Xylene	10.640	106	450386	143.559	ug/l	98
70) Styrene	10.653	104	777762	153.173	ug/l	100
71) Bromoform	10.799	173	207497	188.187	ug/l #	97
73) Isopropylbenzene	10.964	105	1088275	135.577	ug/l	99
74) N-amyl acetate	10.842	43	575386	142.309	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	402017	130.738	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	335384m	128.467	ug/l	
77) Bromobenzene	11.195	156	285385	139.269	ug/l	98
78) n-propylbenzene	11.305	91	1224689	143.160	ug/l	99
79) 2-Chlorotoluene	11.366	91	773459	134.748	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	914265	137.180	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	145480	170.044	ug/l	91
82) 4-Chlorotoluene	11.451	91	893567	140.669	ug/l	97
83) tert-Butylbenzene	11.713	119	935627	137.660	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	953082	139.821	ug/l	100
85) sec-Butylbenzene	11.890	105	1112980	141.391	ug/l	98
86) p-Isopropyltoluene	12.006	119	958987	147.051	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	521523	141.838	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	522586	140.994	ug/l	99
89) n-Butylbenzene	12.329	91	830990	157.180	ug/l	99
90) Hexachloroethane	12.536	117	179474	170.120	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	534157	140.889	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	94746	144.811	ug/l	92
93) 1,2,4-Trichlorobenzene	13.585	180	354426	159.641	ug/l	99
94) Hexachlorobutadiene	13.725	225	118245	150.492	ug/l	98
95) Naphthalene	13.774	128	1350122	148.198	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	363960	156.686	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043561.D
Acq On : 28 Oct 2024 12:54
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 28 13:31:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:29:33 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043561.D
 Acq On : 28 Oct 2024 12:54
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Quant Time: Oct 28 13:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:29:33 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024

