

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102824

Review By	Mahesh Dadoda	Review On	10/29/2024 12:19:57 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/29/2024 12:21:34 PM
SubDirectory	VX102824	HP Acquire Method	MSVOA_X HP Processing Method 82x102824w.m

STD. NAME	STD REF.#
Tune/Reschk	VP131142
Initial Calibration Stds	VP131143,VP131144,VP131145,VP131147,VP131148,VP131149
CCC	VP131151,VP131152
Internal Standard/PEM	VP128298
ICV/I.BLK	VP131150
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043555.D	28 Oct 2024 10:09		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX043556.D	28 Oct 2024 10:59	Method pass	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX043557.D	28 Oct 2024 11:22	8260D	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX043558.D	28 Oct 2024 11:45		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX043559.D	28 Oct 2024 12:08		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX043560.D	28 Oct 2024 12:31		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX043561.D	28 Oct 2024 12:54		JC/MD	Ok,M
8	VIBLK	VIBLK	VX043562.D	28 Oct 2024 14:04		JC/MD	Ok
9	VSTDICV050	ICVVX102824	VX043563.D	28 Oct 2024 14:27		JC/MD	Ok,M
10	VX1028MBL01	VX1028MBL01	VX043564.D	28 Oct 2024 15:02	Internal standard fail	JC/MD	Not Ok
11	VX1028WBL01	VX1028WBL01	VX043565.D	28 Oct 2024 15:52		JC/MD	Ok
12	VX1028WBS01	VX1028WBS01	VX043566.D	28 Oct 2024 16:28		JC/MD	Ok,M
13	VX1028WBSD01	VX1028WBSD01	VX043567.D	28 Oct 2024 16:51		JC/MD	Ok,M
14	PB164394TB	PB164394TB	VX043568.D	28 Oct 2024 17:14	pH#5.0 A	JC/MD	Ok
15	PB164394ZHE#01	PB164394ZHE#01	VX043569.D	28 Oct 2024 17:38	pH#5.0 A	JC/MD	Ok
16	PB164394ZHE#02	PB164394ZHE#02	VX043570.D	28 Oct 2024 18:01	pH#5.0 A	JC/MD	Ok
17	PB164394ZHE#03	PB164394ZHE#03	VX043571.D	28 Oct 2024 18:24	pH#5.0 A	JC/MD	Ok
18	PB164394ZHE#04	PB164394ZHE#04	VX043572.D	28 Oct 2024 18:47	pH#5.0 A	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102824

Review By	Mahesh Dadoda	Review On	10/29/2024 12:19:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/29/2024 12:21:34 PM		
SubDirectory	VX102824	HP Acquire Method	MSVOA_X	HP Processing Method	82x102824w.m
STD. NAME		STD REF.#			
Tune/Reschk		VP131142			
Initial Calibration Stds		VP131143,VP131144,VP131145,VP131147,VP131148,VP131149			
CCC		VP131151,VP131152			
Internal Standard/PEM		VP128298			
ICV/I.BLK		VP131150			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	PB164394ZHE#05	PB164394ZHE#05	VX043573.D	28 Oct 2024 19:10	pH#5.0 A	JC/MD	Ok
20	PB164394ZHE#06	PB164394ZHE#06	VX043574.D	28 Oct 2024 19:33	pH#5.0 A	JC/MD	Ok
21	PB164394ZHE#07	PB164394ZHE#07	VX043575.D	28 Oct 2024 19:56	pH#5.0 A	JC/MD	Ok
22	PB164394ZHE#08	PB164394ZHE#08	VX043576.D	28 Oct 2024 20:19	pH#5.0 A	JC/MD	Ok
23	PB164394ZHE#09	PB164394ZHE#09	VX043577.D	28 Oct 2024 20:42	pH#5.0 A	JC/MD	Ok
24	PB164394ZHE#10	PB164394ZHE#10	VX043578.D	28 Oct 2024 21:05	pH#5.0 A	JC/MD	Ok
25	PB164394ZHE#11	PB164394ZHE#11	VX043579.D	28 Oct 2024 21:28	pH#5.0 A	JC/MD	Ok
26	PB164394ZHE#12	PB164394ZHE#12	VX043580.D	28 Oct 2024 21:52	pH#5.0 A	JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX043581.D	28 Oct 2024 22:15	out of tune time	JC/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102924

Review By	Semsettin Yesilyurt	Review On	10/30/2024 11:31:42 AM
Supervise By	Maresh Dadoda	Supervise On	10/30/2024 1:24:02 PM
SubDirectory	VX102924	HP Acquire Method	MSVOA_X HP Processing Method 82x102824w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131169
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131171,VP131172 VP128298

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043582.D	29 Oct 2024 08:30		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX043583.D	29 Oct 2024 09:29		JC/MD	Ok,M
3	VX1029MBL01	VX1029MBL01	VX043584.D	29 Oct 2024 10:06		JC/MD	Ok
4	VX1029WBL01	VX1029WBL01	VX043585.D	29 Oct 2024 10:31		JC/MD	Ok
5	VX1029WBS01	VX1029WBS01	VX043586.D	29 Oct 2024 10:57		JC/MD	Ok,M
6	VX1029WBSD01	VX1029WBSD01	VX043587.D	29 Oct 2024 11:20		JC/MD	Ok,M
7	VIBLK	VIBLK	VX043588.D	29 Oct 2024 11:43		JC/MD	Ok
8	P4596-07	TOTE-WATER-COMP	VX043589.D	29 Oct 2024 12:10		JC/MD	Ok,M
9	VIBLK	VIBLK	VX043590.D	29 Oct 2024 12:33		JC/MD	Ok
10	P4578-09	TB-10252024	VX043591.D	29 Oct 2024 12:56	vial A pH<2 TB	JC/MD	Ok
11	P4548-09	TRIP BLANK	VX043592.D	29 Oct 2024 13:19	vial A pH<2 TB	JC/MD	Ok
12	P4548-01	MW-1	VX043593.D	29 Oct 2024 13:43	vial A pH<2	JC/MD	Ok
13	P4548-03	MW-2	VX043594.D	29 Oct 2024 14:06	vial A pH<2	JC/MD	Ok
14	P4548-05	MW-3	VX043595.D	29 Oct 2024 14:29	vial A pH<2	JC/MD	Ok
15	P4548-07	MW-4	VX043596.D	29 Oct 2024 14:52	vial A pH<2	JC/MD	Ok
16	P4578-01	WB-305-SW	VX043597.D	29 Oct 2024 15:15	vial A pH<2	JC/MD	Ok
17	P4578-02	WB-305-SW-1	VX043598.D	29 Oct 2024 15:38	vial A pH<2	JC/MD	Ok
18	P4600-01	BP-VPB-190-TB-20241	VX043599.D	29 Oct 2024 16:01	vial A pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102924

Review By	Semsettin Yesilyurt	Review On	10/30/2024 11:31:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	10/30/2024 1:24:02 PM		
SubDirectory	VX102924	HP Acquire Method	MSVOA_X	HP Processing Method	82x102824w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131169			
CCC		VP131171,VP131172			
Internal Standard/PEM		VP128298			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4600-02	BP-VPB-190-DUP-2024	VX043600.D	29 Oct 2024 16:24	vial A pH<2	JC/MD	Ok
20	P4600-03	BP-VPB-190-GW-438-4	VX043601.D	29 Oct 2024 16:48	vial A pH<2	JC/MD	Ok
21	P4600-04	BP-VPB-190-GW-458-4	VX043602.D	29 Oct 2024 17:11	vial A pH<2	JC/MD	Ok
22	P4600-05	BP-VPB-190-GW-478-4	VX043603.D	29 Oct 2024 17:34	vial A pH<2 Internal standard fail	JC/MD	ReRun
23	P4600-06	BP-VPB-190-GW-498-5	VX043604.D	29 Oct 2024 17:57	vial A pH<2	JC/MD	Ok
24	P4600-07	BP-VPB-190-GW-518-5	VX043605.D	29 Oct 2024 18:20	vial A pH<2	JC/MD	Ok
25	P4562-01	Storage-Blank-SOIL-RE	VX043606.D	29 Oct 2024 18:43	vial A pH<2	JC/MD	Ok
26	P4562-01	Storage-Blank-SOIL-RE	VX043607.D	29 Oct 2024 19:06	vial A pH<2 Hit of com# 3	JC/MD	ReRun
27	VSTDCCC050	VSTDCCC050EC	VX043608.D	29 Oct 2024 19:29		JC/MD	Ok,M

M : Manual Integration