


LB133167

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	C/D
0.0	0.000	-0.080	-0.080	-18	0.000	-18	0				0/0
0.2	0.200	0.183	-0.017	825	0.0 %	825					-9
2.5	2.500	2.458	-0.042	8133	0.0 %	8133					-2
5.0	5.000	5.160	0.160	16813	0.0 %	16813					3
7.5	7.500	7.767	0.267	25186	0.0 %	25186					4
10.0	10.000	9.712	-0.288	31433	0.0 %	31433					-3

LB133167 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Oct 2024 10:41:11

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.00007470A	PPB	28 Oct 2024 10:44:16	S	Std
0.2 - 1	50.2	825	-	0.20007470A	PPB	28 Oct 2024 10:51:28	S	Std
2.5 - 1	52.5	8133	-	2.50007470A	PPB	28 Oct 2024 10:53:45	S	Std
5.0 - 1	55	16813	-	5.00007470A	PPB	28 Oct 2024 10:56:02	S	Std
7.5 - 1	57.5	25186	-	7.50007470A	PPB	28 Oct 2024 11:01:03	S	Std
10.0 - 1	510	31433	-	10.00007470A	PPB	28 Oct 2024 11:05:55	S	Std
ICV55 - 1	ICV55	12127	3.7014	-7470A	PPB	28 Oct 2024 11:09:29	U	SMPL
ICB55 - 1	ICB55	-64	-0.0939	-7470A	PPB	28 Oct 2024 11:14:18	U	SMPL
CCV79 - 1	CCV79	17032	5.2284	-7470A	PPB	28 Oct 2024 11:16:32	U	SMPL
CCB79 - 1	CCB79	86	-0.0472	-7470A	PPB	28 Oct 2024 11:21:22	U	SMPL
CRA - 1	CRA	804	0.1763	-7470A	PPB	28 Oct 2024 11:23:37	U	SMPL
HighStd - 1	HighStd	30915	9.5505	-7470A	PPB	28 Oct 2024 11:25:51	U	SMPL
ChkStd - 1	ChkStd	22638	6.9737	-7470A	PPB	28 Oct 2024 11:28:07	U	SMPL
PB164472BL - 1	PBW	16	-0.0690	-7470A	PPB	28 Oct 2024 11:32:59	U	SMPL
PB164472BS - 1	LCSW	12719	3.8857	-7470A	PPB	28 Oct 2024 11:35:15	U	SMPL
P4549-02 - 1	TT-067-IDWGW-20241024	-169	-0.1266	-7470A	PPB	28 Oct 2024 11:37:31	U	SMPL
P4549-02DUP - 1	TT-067-IDWGW-20241024DUP	124	-0.0354	-7470A	PPB	28 Oct 2024 11:39:48	U	SMPL
P4549-02MS - 1	TT-067-IDWGW-20241024MS	13399	4.0974	-7470A	PPB	28 Oct 2024 11:42:06	U	SMPL
P4549-02MSD - 1	TT-067-IDWGW-20241024MSD	12159	3.7113	-7470A	PPB	28 Oct 2024 11:44:24	U	SMPL
P4549-03 - 1	TT-068-IDWGW-20241024	394	0.0488	-7470A	PPB	28 Oct 2024 11:46:41	U	SMPL
CCV80 - 1	CCV80	16123	4.9454	-7470A	PPB	28 Oct 2024 11:48:58	U	SMPL
CCB80 - 1	CCB80	117	-0.0376	-7470A	PPB	28 Oct 2024 11:53:45	U	SMPL
P4549-04 - 1	TT-069-IDWGW-20241024	76	-0.0504	-7470A	PPB	28 Oct 2024 11:56:03	U	SMPL
P4564-01 - 1	100324	17	-0.0687	-7470A	PPB	28 Oct 2024 11:58:18	U	SMPL
P4578-01 - 1	WB-305-SW	-186	-0.1319	-7470A	PPB	28 Oct 2024 12:00:33	U	SMPL
P4578-02 - 1	WB-305-SW-1	24	-0.0665	-7470A	PPB	28 Oct 2024 12:02:49	U	SMPL
P4549-02LX5 - 1		-24	-0.0815	-7470A	PPB	28 Oct 2024 12:05:05	U	SMPL
P4549-02A - 1		12725	3.8875	-7470A	PPB	28 Oct 2024 12:07:21	U	SMPL
CCV81 - 1	CCV81	16246	4.9837	-7470A	PPB	28 Oct 2024 12:09:37	U	SMPL
CCB81 - 1	CCB81	79	-0.0494	-7470A	PPB	28 Oct 2024 12:14:26	U	SMPL