



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** P4549
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4549 **SAS No.:** P4549
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	88	40 - 160	CV	10/28/2024	11:23	LB133167
CRI01	Aluminum	102	100	102	40 - 160	P	10/28/2024	13:12	LB133175
	Antimony	50.0	50.0	100	40 - 160	P	10/28/2024	13:12	LB133175
	Arsenic	19.1	20.0	96	40 - 160	P	10/28/2024	13:12	LB133175
	Barium	93.7	100	94	40 - 160	P	10/28/2024	13:12	LB133175
	Beryllium	5.64	6.0	94	40 - 160	P	10/28/2024	13:12	LB133175
	Cadmium	5.55	6.0	92	40 - 160	P	10/28/2024	13:12	LB133175
	Calcium	1930	2000	96	40 - 160	P	10/28/2024	13:12	LB133175
	Chromium	9.13	10.0	91	40 - 160	P	10/28/2024	13:12	LB133175
	Cobalt	29.3	30.0	98	40 - 160	P	10/28/2024	13:12	LB133175
	Copper	21.1	20.0	105	40 - 160	P	10/28/2024	13:12	LB133175
	Iron	98.8	100	99	40 - 160	P	10/28/2024	13:12	LB133175
	Lead	11.5	12.0	96	40 - 160	P	10/28/2024	13:12	LB133175
	Magnesium	1900	2000	95	40 - 160	P	10/28/2024	13:12	LB133175
	Manganese	19.1	20.0	95	40 - 160	P	10/28/2024	13:12	LB133175
	Nickel	39.2	40.0	98	40 - 160	P	10/28/2024	13:12	LB133175
	Potassium	1940	2000	97	40 - 160	P	10/28/2024	13:12	LB133175
	Selenium	17.6	20.0	88	40 - 160	P	10/28/2024	13:12	LB133175
	Silver	11.0	10.0	110	40 - 160	P	10/28/2024	13:12	LB133175
	Sodium	1780	2000	89	40 - 160	P	10/28/2024	13:12	LB133175
	Thallium	34.9	40.0	87	40 - 160	P	10/28/2024	13:12	LB133175
	Vanadium	36.9	40.0	92	40 - 160	P	10/28/2024	13:12	LB133175
	Zinc	37.3	40.0	93	40 - 160	P	10/28/2024	13:12	LB133175