



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tetra Tech NUS, Inc.				SDG No.:	P4549				
Contract:	TETR06	Lab Code:	CHEM	Case No.:	P4549	SAS No.:	P4549			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB55	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	10/28/2024	11:14	LB133167

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P4549

Contract: TETR06

Lab Code: CHEM

Case No.: P4549

SAS No.: P4549

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	10/28/2024	11:21	LB133167
CCB80	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	10/28/2024	11:53	LB133167
CCB81	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	10/28/2024	12:14	LB133167

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. SDG No.: P4549
Contract: TETR06 Lab Code: CHEM Case No.: P4549 SAS No.: P4549

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	13:08	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	13:08	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:08	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	13:08	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	13:08	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	13:08	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	13:08	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	13:08	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	13:08	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:08	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	13:08	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	13:08	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	13:08	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	13:08	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	13:08	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	13:08	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:08	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	13:08	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	13:08	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	13:08	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	13:08	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	13:08	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.		SDG No.: P4549								
Contract: TETR06	Lab Code: CHEM	Case No.: P4549	SAS No.: P4549							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	13:39	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	13:39	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:39	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	13:39	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	13:39	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	13:39	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	13:39	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	13:39	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	13:39	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:39	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	13:39	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	13:39	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	13:39	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	13:39	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	13:39	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	13:39	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	13:39	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	13:39	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	13:39	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	13:39	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	13:39	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	13:39	LB133175
CCB02	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	14:30	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	14:30	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	14:30	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	14:30	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	14:30	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	14:30	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	14:30	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	14:30	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	14:30	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	14:30	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	14:30	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	14:30	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	14:30	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	14:30	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	14:30	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	14:30	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	14:30	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.		SDG No.: P4549								
Contract: TETR06	Lab Code: CHEM	Case No.: P4549	SAS No.: P4549							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	14:30	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	14:30	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	14:30	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	14:30	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	14:30	LB133175
CCB03	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	15:20	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	15:20	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	15:20	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	15:20	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	15:20	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	15:20	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	15:20	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	15:20	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	15:20	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	15:20	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	15:20	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	15:20	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	15:20	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	15:20	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	15:20	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	15:20	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	15:20	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	15:20	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	15:20	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	15:20	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	15:20	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	15:20	LB133175
CCB04	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	16:10	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	16:10	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	16:10	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	16:10	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	16:10	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	16:10	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	16:10	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	16:10	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	16:10	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	16:10	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	16:10	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	16:10	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.		SDG No.: P4549								
Contract: TETR06	Lab Code: CHEM	Case No.: P4549	SAS No.: P4549							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	16:10	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	16:10	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	16:10	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	16:10	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	16:10	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	16:10	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	16:10	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	16:10	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	16:10	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	16:10	LB133175
CCB05	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	17:01	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	17:01	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:01	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	17:01	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	17:01	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	17:01	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	17:01	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	17:01	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	17:01	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:01	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	17:01	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	17:01	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	17:01	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	17:01	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	17:01	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	17:01	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:01	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	17:01	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	17:01	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	17:01	LB133175
CCB06	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	17:01	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	17:01	LB133175
	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	17:54	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	17:54	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:54	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	17:54	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	17:54	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	17:54	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	17:54	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P4549

Contract: TETR06

Lab Code: CHEM

Case No.: P4549

SAS No.: P4549

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	17:54	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	17:54	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:54	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	17:54	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	17:54	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	17:54	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	17:54	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	17:54	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	17:54	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	17:54	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	17:54	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	17:54	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	17:54	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	17:54	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	17:54	LB133175
CCB07	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	18:47	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	18:47	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	18:47	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	18:47	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	18:47	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	18:47	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	18:47	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	18:47	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	18:47	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	18:47	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	18:47	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	18:47	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	18:47	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	18:47	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	18:47	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	18:47	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	18:47	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	18:47	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	18:47	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	18:47	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	18:47	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	18:47	LB133175
CCB08	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	19:37	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	19:37	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.		SDG No.: P4549								
Contract: TETR06	Lab Code: CHEM	Case No.: P4549	SAS No.: P4549							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	19:37	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	19:37	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	19:37	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	19:37	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	19:37	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	19:37	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	19:37	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	19:37	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	19:37	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	19:37	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	19:37	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	19:37	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	19:37	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	19:37	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	19:37	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	19:37	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	19:37	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	19:37	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	19:37	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	19:37	LB133175
CCB09	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	20:26	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	20:26	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	20:26	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	20:26	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	20:26	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	20:26	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	20:26	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	20:26	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	20:26	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	20:26	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	20:26	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	20:26	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	20:26	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	20:26	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	20:26	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	20:26	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	20:26	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	20:26	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	20:26	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P4549

Contract: TETR06

Lab Code: CHEM

Case No.: P4549

SAS No.: P4549

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	20:26	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	20:26	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	20:26	LB133175
CCB10	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	21:25	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	21:25	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:25	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	21:25	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	21:25	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	21:25	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	21:25	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	21:25	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	21:25	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:25	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	21:25	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	21:25	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	21:25	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	21:25	LB133175
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	21:25	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	21:25	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:25	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	21:25	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	21:25	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	21:25	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	21:25	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	21:25	LB133175
	Aluminum	100	+/-100	U	80.0	100	P	10/28/2024	21:43	LB133175
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	10/28/2024	21:43	LB133175
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:43	LB133175
	Barium	100	+/-100	U	25.0	100	P	10/28/2024	21:43	LB133175
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	21:43	LB133175
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	10/28/2024	21:43	LB133175
	Calcium	2000	+/-2000	U	500	2000	P	10/28/2024	21:43	LB133175
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	21:43	LB133175
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	10/28/2024	21:43	LB133175
	Copper	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:43	LB133175
	Iron	100	+/-100	U	80.0	100	P	10/28/2024	21:43	LB133175
	Lead	12.0	+/-12.0	U	9.60	12.0	P	10/28/2024	21:43	LB133175
	Magnesium	2000	+/-2000	U	500	2000	P	10/28/2024	21:43	LB133175
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	10/28/2024	21:43	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech NUS, Inc.</u>		SDG No.: <u>P4549</u>								
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P4549</u>	SAS No.: <u>P4549</u>				
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	21:43	LB133175
	Potassium	2000	+/-2000	U	1600	2000	P	10/28/2024	21:43	LB133175
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	10/28/2024	21:43	LB133175
	Silver	10.0	+/-10.0	U	5.00	10.0	P	10/28/2024	21:43	LB133175
	Sodium	2000	+/-2000	U	1000	2000	P	10/28/2024	21:43	LB133175
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	21:43	LB133175
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	10/28/2024	21:43	LB133175
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	10/28/2024	21:43	LB133175