

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>P4549</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P4549</u>
Matrix:	<u>Water</u>	Sample ID:	<u>P4549-02</u>	Client ID:	<u>TT-067-IDWGW-20241024DUP</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>P4549-02DUP</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4549
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4549 **SAS No.:** P4549
Matrix: Water **Sample ID:** P4549-02MS **Client ID:** TT-067-IDWGW-20241024MSD
Percent Solids for Sample: NA **Duplicate ID** P4549-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.10		3.71		10		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4549
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4549 **SAS No.:** P4549
Matrix: Water **Sample ID:** P4549-04 **Client ID:** TT-069-IDWGW-20241024DUP
Percent Solids for Sample: NA **Duplicate ID** P4549-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	13000		13100		1		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	96.5		96.2		0		P
Beryllium	ug/L	20	1.11	J	1.11	J	0		P
Cadmium	ug/L	20	0.50	J	0.51	J	3		P
Calcium	ug/L	20	9080		9070		0		P
Chromium	ug/L	20	15.1		15.0		1		P
Cobalt	ug/L	20	5.18	J	5.21	J	1		P
Copper	ug/L	20	11.2		11.0		2		P
Iron	ug/L	20	7460		7700		3		P
Lead	ug/L	20	16.3		15.9		2		P
Magnesium	ug/L	20	3380		3400		1		P
Manganese	ug/L	20	192		191		1		P
Nickel	ug/L	20	9.45	J	9.28	J	2		P
Potassium	ug/L	20	1480		1510		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	43600		45100		3		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	13.5	J	13.1	J	3		P
Zinc	ug/L	20	35.7		32.4		10		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4549
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4549 **SAS No.:** P4549
Matrix: Water **Sample ID:** P4549-04MS **Client ID:** TT-069-IDWGW-20241024MSD
Percent Solids for Sample: NA **Duplicate ID** P4549-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	19900		19800		1		P
Antimony	ug/L	20	117		116		1		P
Arsenic	ug/L	20	343		342		0		P
Barium	ug/L	20	185		184		1		P
Beryllium	ug/L	20	85.6		86.6		1		P
Cadmium	ug/L	20	86.9		86.6		0		P
Calcium	ug/L	20	9530		9550		0		P
Chromium	ug/L	20	199		204		2		P
Cobalt	ug/L	20	98.6		98.6		0		P
Copper	ug/L	20	146		145		1		P
Iron	ug/L	20	9820		10000		2		P
Lead	ug/L	20	428		428		0		P
Magnesium	ug/L	20	4690		4730		1		P
Manganese	ug/L	20	279		280		0		P
Nickel	ug/L	20	242		241		0		P
Potassium	ug/L	20	5960		6060		2		P
Selenium	ug/L	20	824		821		0		P
Silver	ug/L	20	30.3		30.8		2		P
Sodium	ug/L	20	44500		45200		2		P
Thallium	ug/L	20	777		735		6		P
Vanadium	ug/L	20	145		145		0		P
Zinc	ug/L	20	130		131		1		P