

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

Continuing Calib Date: 10/25/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:43 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.38	6.38	6.28	6.48	0.00
Aroclor-1254-2	(2)	6.59	6.59	6.49	6.69	0.00
Aroclor-1254-3	(3)	6.96	6.96	6.86	7.06	0.00
Aroclor-1254-4	(4)	7.24	7.24	7.14	7.34	0.00
Aroclor-1254-5	(5)	7.66	7.66	7.56	7.76	0.00
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.06	10.06	9.96	10.16	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

Continuing Calib Date: 10/25/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:43 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1254-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1254-3	(3)	6.08	6.08	5.98	6.18	0.00
Aroclor-1254-4	(4)	6.30	6.30	6.20	6.40	0.00
Aroclor-1254-5	(5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene		3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00