

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

Continuing Calib Date: 10/28/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:28 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1221-1 (1)   | 4.57       | 4.57      | 4.47      | 4.67  | 0.00       |
| Aroclor-1221-2 (2)   | 4.66       | 4.66      | 4.56      | 4.76  | 0.00       |
| Aroclor-1221-3 (3)   | 4.73       | 4.73      | 4.63      | 4.83  | 0.00       |
| Tetrachloro-m-xylene | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   | 10.06      | 10.06     | 9.96      | 10.16 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

Continuing Calib Date: 10/28/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 16:28 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1221-1 (1)   | 3.86       | 3.86      | 3.76      | 3.96 | 0.00       |
| Aroclor-1221-2 (2)   | 3.94       | 3.94      | 3.84      | 4.04 | 0.00       |
| Aroclor-1221-3 (3)   | 4.02       | 4.02      | 3.92      | 4.12 | 0.00       |
| Tetrachloro-m-xylene | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   | 8.63       | 8.64      | 8.54      | 8.74 | 0.01       |