

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107382.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	541.940	500.000	8.4
Aroclor-1016-2	5.544	5.445	5.645	534.490	500.000	6.9
Aroclor-1016-3	5.606	5.507	5.707	524.670	500.000	4.9
Aroclor-1016-4	5.704	5.604	5.804	550.740	500.000	10.1
Aroclor-1016-5	5.998	5.899	6.099	549.630	500.000	9.9
Aroclor-1260-1	7.125	7.026	7.226	524.670	500.000	4.9
Aroclor-1260-2	7.382	7.282	7.482	536.310	500.000	7.3
Aroclor-1260-3	7.743	7.643	7.843	515.030	500.000	3.0
Aroclor-1260-4	7.968	7.868	8.068	522.860	500.000	4.6
Aroclor-1260-5	8.281	8.181	8.381	527.620	500.000	5.5
Decachlorobiphenyl	10.061	9.958	10.158	52.130	50.000	4.3
Tetrachloro-m-xylene	4.373	4.274	4.474	55.320	50.000	10.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107382.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.626	4.826	553.460	500.000	10.7
Aroclor-1016-2	4.745	4.645	4.845	567.100	500.000	13.4
Aroclor-1016-3	4.921	4.821	5.021	533.360	500.000	6.7
Aroclor-1016-4	4.962	4.862	5.062	510.120	500.000	2.0
Aroclor-1016-5	5.175	5.076	5.276	544.660	500.000	8.9
Aroclor-1260-1	6.206	6.107	6.307	548.980	500.000	9.8
Aroclor-1260-2	6.393	6.294	6.494	584.550	500.000	16.9
Aroclor-1260-3	6.547	6.448	6.648	558.400	500.000	11.7
Aroclor-1260-4	7.017	6.918	7.118	558.220	500.000	11.6
Aroclor-1260-5	7.257	7.159	7.359	591.990	500.000	18.4
Decachlorobiphenyl	8.638	8.539	8.739	53.990	50.000	8.0
Tetrachloro-m-xylene	3.645	3.545	3.745	55.010	50.000	10.0