

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO107409.D Time Analyzed: 00:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	496.520	500.000	-0.7
Aroclor-1016-2	5.544	5.445	5.645	485.870	500.000	-2.8
Aroclor-1016-3	5.606	5.507	5.707	472.730	500.000	-5.5
Aroclor-1016-4	5.702	5.604	5.804	493.570	500.000	-1.3
Aroclor-1016-5	5.997	5.899	6.099	497.580	500.000	-0.5
Aroclor-1260-1	7.125	7.026	7.226	465.740	500.000	-6.9
Aroclor-1260-2	7.381	7.282	7.482	492.130	500.000	-1.6
Aroclor-1260-3	7.743	7.643	7.843	486.130	500.000	-2.8
Aroclor-1260-4	7.967	7.868	8.068	492.230	500.000	-1.6
Aroclor-1260-5	8.280	8.181	8.381	496.360	500.000	-0.7
Decachlorobiphenyl	10.060	9.958	10.158	46.720	50.000	-6.6
Tetrachloro-m-xylene	4.372	4.274	4.474	50.430	50.000	0.9

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO107409.D Time Analyzed: 00:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	527.800	500.000	5.6
Aroclor-1016-2	4.744	4.645	4.845	528.560	500.000	5.7
Aroclor-1016-3	4.920	4.821	5.021	503.880	500.000	0.8
Aroclor-1016-4	4.962	4.862	5.062	478.610	500.000	-4.3
Aroclor-1016-5	5.175	5.076	5.276	507.540	500.000	1.5
Aroclor-1260-1	6.206	6.107	6.307	518.160	500.000	3.6
Aroclor-1260-2	6.393	6.294	6.494	532.900	500.000	6.6
Aroclor-1260-3	6.547	6.448	6.648	524.730	500.000	4.9
Aroclor-1260-4	7.017	6.918	7.118	533.380	500.000	6.7
Aroclor-1260-5	7.259	7.159	7.359	554.010	500.000	10.8
Decachlorobiphenyl	8.638	8.539	8.739	46.430	50.000	-7.1
Tetrachloro-m-xylene	3.644	3.545	3.745	51.580	50.000	3.2