

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107421.D Time Analyzed: 11:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.519	5.423	5.623	467.400	500.000	-6.5
Aroclor-1016-2	5.542	5.445	5.645	458.050	500.000	-8.4
Aroclor-1016-3	5.604	5.507	5.707	455.130	500.000	-9.0
Aroclor-1016-4	5.701	5.604	5.804	473.820	500.000	-5.2
Aroclor-1016-5	5.996	5.899	6.099	478.100	500.000	-4.4
Aroclor-1260-1	7.124	7.026	7.226	463.460	500.000	-7.3
Aroclor-1260-2	7.379	7.282	7.482	471.410	500.000	-5.7
Aroclor-1260-3	7.741	7.643	7.843	454.480	500.000	-9.1
Aroclor-1260-4	7.965	7.868	8.068	438.720	500.000	-12.3
Aroclor-1260-5	8.278	8.181	8.381	474.790	500.000	-5.0
Decachlorobiphenyl	10.054	9.958	10.158	46.260	50.000	-7.5
Tetrachloro-m-xylene	4.371	4.274	4.474	47.030	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107421.D Time Analyzed: 11:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.626	4.826	512.280	500.000	2.5
Aroclor-1016-2	4.743	4.645	4.845	516.410	500.000	3.3
Aroclor-1016-3	4.919	4.821	5.021	495.570	500.000	-0.9
Aroclor-1016-4	4.960	4.862	5.062	477.890	500.000	-4.4
Aroclor-1016-5	5.173	5.076	5.276	501.090	500.000	0.2
Aroclor-1260-1	6.204	6.107	6.307	494.270	500.000	-1.1
Aroclor-1260-2	6.391	6.294	6.494	526.580	500.000	5.3
Aroclor-1260-3	6.544	6.448	6.648	502.020	500.000	0.4
Aroclor-1260-4	7.015	6.918	7.118	495.390	500.000	-0.9
Aroclor-1260-5	7.255	7.159	7.359	527.980	500.000	5.6
Decachlorobiphenyl	8.634	8.539	8.739	47.340	50.000	-5.3
Tetrachloro-m-xylene	3.643	3.545	3.745	49.980	50.000	0.0