

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101524\
 Data File : P0107210.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2024 02:18
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 04:21:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 16 02:38:17 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.373	3.644	237.4E6	83283643	25.493	24.380
2) SA Decachlor...	10.058	8.638	110.0E6	122.3E6	25.930	25.345
Target Compounds						
41) L9 AR-1268-1	8.588	7.541	95440132	102.7E6	261.750	248.179
42) L9 AR-1268-2	8.679	7.606	84909424	94674968	261.370	246.859
43) L9 AR-1268-3	8.908	7.812	73384675	84688940	259.658	249.426
44) L9 AR-1268-4	9.314	8.097	29107021	32315199	253.734	253.880
45) L9 AR-1268-5	9.721	8.386	216.1E6	249.1E6	255.104	244.896



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101524\
Data File : PO107210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2024 02:18
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 04:21:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 16 02:38:17 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.373 min
Delta R.T.: 0.000 min
Response: 237379747
Conc: 25.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: 0.000 min
Response: 83283643
Conc: 24.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.000 min
Response: 109995659
Conc: 25.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.638 min
Delta R.T.: 0.000 min
Response: 122251099
Conc: 25.34 ng/ml



#41 AR-1268-1


R.T.: 8.588 min
Delta R.T.: 0.002 min
Response: 95440132
Conc: 261.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.541 min
Delta R.T.: 0.000 min
Response: 102712368
Conc: 248.18 ng/ml




#42 AR-1268-2

R.T.: 8.679 min
Delta R.T.: 0.001 min
Response: 84909424
Conc: 261.37 ng/ml



#42 AR-1268-2


R.T.: 7.606 min
Delta R.T.: 0.000 min
Response: 94674968
Conc: 246.86 ng/ml



#43 AR-1268-3


R.T.: 8.908 min
Delta R.T.: 0.000 min
Response: 73384675
Conc: 259.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.812 min
Delta R.T.: 0.000 min
Response: 84688940
Conc: 249.43 ng/ml




#44 AR-1268-4

R.T.: 9.314 min
Delta R.T.: 0.002 min
Response: 29107021
Conc: 253.73 ng/ml



#44 AR-1268-4


R.T.: 8.097 min
Delta R.T.: 0.000 min
Response: 32315199
Conc: 253.88 ng/ml



#45 AR-1268-5

R.T.: 9.721 min
Delta R.T.: 0.000 min
Response: 216089361
Conc: 255.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.386 min
Delta R.T.: 0.000 min
Response: 249078869
Conc: 244.90 ng/ml