

Surrogate Summary

SDG No.: **P4549**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	22.8	114		60	140
I.BLK-PO107386.D	PIBLK-PO107386.D	Tetrachloro-m-xylene	1	20	22.3	112		60	140
		Decachlorobiphenyl	1	20	22.6	113		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	22.8	114		60	140
P4549-04	TT-069-IDWGW-20241024	Tetrachloro-m-xylene	1	20	17.2	86		35	137
		Decachlorobiphenyl	1	20	6.05	30	*	40	135
		Tetrachloro-m-xylene	2	20	19.2	96		35	137
		Decachlorobiphenyl	2	20	5.68	28	*	40	135
PB164403BS	PB164403BS	Tetrachloro-m-xylene	1	20	20.7	104		35	137
		Decachlorobiphenyl	1	20	21.0	105		40	135
		Tetrachloro-m-xylene	2	20	20.4	102		35	137
		Decachlorobiphenyl	2	20	18.8	94		40	135
PB164403BSD	PB164403BSD	Tetrachloro-m-xylene	1	20	21.5	108		35	137
		Decachlorobiphenyl	1	20	21.1	105		40	135
		Tetrachloro-m-xylene	2	20	20.5	103		35	137
		Decachlorobiphenyl	2	20	19.4	97		40	135
I.BLK-PO107401.D	PIBLK-PO107401.D	Tetrachloro-m-xylene	1	20	22.1	111		60	140
		Decachlorobiphenyl	1	20	21.2	106		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
P4549-02	TT-067-IDWGW-20241024	Tetrachloro-m-xylene	1	20	18.7	93		35	137
		Decachlorobiphenyl	1	20	6.81	34	*	40	135
		Tetrachloro-m-xylene	2	20	19.4	97		35	137
		Decachlorobiphenyl	2	20	6.39	32	*	40	135
P4549-03	TT-068-IDWGW-20241024	Tetrachloro-m-xylene	1	20	17.9	89		35	137
		Decachlorobiphenyl	1	20	5.86	29	*	40	135
		Tetrachloro-m-xylene	2	20	18.4	92		35	137
		Decachlorobiphenyl	2	20	5.62	28	*	40	135
I.BLK-PO107413.D	PIBLK-PO107413.D	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	21.4	107		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	20.4	102		60	140
I.BLK-PO107424.D	PIBLK-PO107424.D	Tetrachloro-m-xylene	1	20	19.3	96		60	140
		Decachlorobiphenyl	1	20	20.2	101		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	19.6	98		60	140
PB164403BL	PB164403BL	Tetrachloro-m-xylene	1	20	19.6	98		35	137

Surrogate Summary

SDG No.: P4549

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
PB164403BL	PB164403BL	Decachlorobiphenyl	1	20	22.5	113		40	135
		Tetrachloro-m-xylene	2	20	20.6	103		35	137
		Decachlorobiphenyl	2	20	22.6	113		40	135
P4549-02RE	TT-067-IDWGW-20241024RE	Tetrachloro-m-xylene	1	20	18.7	94		35	137
		Decachlorobiphenyl	1	20	7.75	39	*	40	135
		Tetrachloro-m-xylene	2	20	19.7	98		35	137
		Decachlorobiphenyl	2	20	7.36	37	*	40	135
		Tetrachloro-m-xylene	1	20	17.5	87		35	137
P4549-03RE	TT-068-IDWGW-20241024RE	Decachlorobiphenyl	1	20	6.58	33	*	40	135
		Tetrachloro-m-xylene	2	20	17.9	90		35	137
		Decachlorobiphenyl	2	20	6.42	32	*	40	135
		Tetrachloro-m-xylene	1	20	18.0	90		35	137
		Decachlorobiphenyl	1	20	6.75	34	*	40	135
P4549-04RE	TT-069-IDWGW-20241024RE	Tetrachloro-m-xylene	2	20	18.8	94		35	137
		Decachlorobiphenyl	2	20	6.42	32	*	40	135
		Tetrachloro-m-xylene	1	20	19.9	100		60	140
		Decachlorobiphenyl	1	20	20.5	103		60	140
		Tetrachloro-m-xylene	2	20	20.5	103		60	140
I.BLK-PO107440.D	PIBLK-PO107440.D	Decachlorobiphenyl	2	20	20.4	102		60	140