

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
 Data File : VN084504.D
 Acq On : 25 Oct 2024 10:04
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 26 05:24:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:22:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 M	Dichlorodifluoromethane	1.832	1.676	8.5	89	0.00
3 M	Chloromethane	2.084	1.831	12.1	87	0.00
4 M	Vinyl Chloride	2.044	1.874	8.3	91	0.00
5 M	Bromomethane	1.354	1.189	12.2	86	0.00
6 M	Chloroethane	1.340	1.274	4.9	93	0.00
7 M	Trichlorofluoromethane	3.329	3.123	6.2	94	0.00
8 T	Diethyl Ether	1.224	1.151	6.0	96	0.00
9	1,1,2-Trichlorotrifluoroeth	1.872	1.862	0.5	101	0.01
10 M	1,1-Dichloroethene	1.868	1.715	8.2	96	0.00
11	Methyl Iodide	2.455	2.262	7.9	94	0.00
12	Methyl Acetate	2.339	2.483	-6.2	111	0.00
13 M	Acrolein	0.434	0.431	0.7	110	0.00
14 M	Acrylonitrile	1.051	0.953	9.3	91	0.00
15 M	Acetone	0.300	0.368	-22.7	132	0.00
16 M	Carbon Disulfide	5.562	4.632	16.7	83	0.00
17	Allyl chloride	3.028	2.791	7.8	92	0.00
18 M	Methylene Chloride	2.104	1.939	7.8	92	0.00
19 M	trans-1,2-Dichloroethene	1.977	1.797	9.1	91	0.00
20 T	Diisopropyl ether	6.510	6.189	4.9	95	0.00
21 M	1,1-Dichloroethane	3.736	3.621	3.1	97	0.00
22 M	cis-1,2-Dichloroethene	2.300	2.240	2.6	97	0.00
23 M	tert-Butyl Alcohol	0.403	0.364	9.7	93	0.00
24 M	Methyl tert-Butyl Ether	6.242	5.963	4.5	97	-0.01
25 M	Chloroform	3.783	3.824	-1.1	100	0.00
26	Cyclohexane	3.124	2.756	11.8	86	0.00
27 s	1,2-Dichloroethane-d4	2.415	2.412	0.1	100	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
29	1,1-Dichloropropene	0.492	0.447	9.1	94	0.00
30 M	2-Butanone	0.262	0.252	3.8	102	0.00
31	2,2-Dichloropropane	0.579	0.569	1.7	102	0.00
32 M	1,1,1-Trichloroethane	0.629	0.603	4.1	96	0.00
33 M	Carbon Tetrachloride	0.545	0.504	7.5	96	0.00
34 M	Benzene	1.542	1.442	6.5	96	0.00
35	Methacrylonitrile	0.280	0.255	8.9	96	0.00
36 M	1,2-Dichloroethane	0.514	0.488	5.1	97	0.00
37 M	Trichloroethene	0.355	0.335	5.6	95	0.00
38	Methylcyclohexane	0.541	0.486	10.2	97	0.00
39 M	1,2-Dichloropropane	0.376	0.353	6.1	95	0.00
40	Dibromomethane	0.263	0.248	5.7	98	0.00
41 M	Bromodichloromethane	0.558	0.524	6.1	97	0.00
42 M	Vinyl Acetate	0.912	0.805	11.7	94	0.00
43	Ethyl Acetate	0.517	0.440	14.9	89	0.00
44	Isopropyl Acetate	0.846	0.973	-15.0	120	0.00
45 T	1,4-Dioxane	0.008	0.007	12.5	97	0.00
46	Methyl methacrylate	0.409	0.351	14.2	90	0.00
47	n-amyl Acetate	0.753	0.668	11.3	94	0.00
48 M	t-1,3-Dichloropropene	0.576	0.532	7.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102524\
 Data File : VN084504.D
 Acq On : 25 Oct 2024 10:04
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 26 05:24:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:22:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.604	0.565	6.5	95	0.00
50 M	1,1,2-Trichloroethane	0.353	0.335	5.1	98	0.00
51	Ethyl methacrylate	0.583	0.530	9.1	96	0.00
52	1,3-Dichloropropane	0.617	0.579	6.2	96	0.00
53 M	Dibromochloromethane	0.417	0.392	6.0	98	0.00
54 M	1,2-Dibromoethane	0.364	0.334	8.2	93	0.00
55 M	2-Chloroethyl vinyl ether	0.265	0.234	11.7	97	0.00
56 M	Bromoform	0.277	0.257	7.2	101	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
58 M	4-Methyl-2-Pentanone	0.559	0.530	5.2	95	0.00
59 M	2-Hexanone	0.415	0.406	2.2	100	0.00
60 S	4-Bromofluorobenzene	0.495	0.511	-3.2	104	0.00
61 M	Tetrachloroethene	0.339	0.348	-2.7	102	0.00
62 M	Toluene	1.756	1.683	4.2	96	0.00
63 S	Toluene-d8	1.390	1.390	0.0	100	0.00
64 M	Chlorobenzene	1.093	1.036	5.2	95	0.00
65	1,1,1,2-Tetrachloroethane	0.380	0.376	1.1	98	0.00
66 M	Ethyl Benzene	1.908	1.839	3.6	97	0.00
67 M	m/p-Xylenes	0.724	0.700	3.3	97	0.00
68 M	o-Xylene	0.698	0.660	5.4	95	0.00
69 M	Styrene	1.198	1.135	5.3	96	0.00
70	Isopropylbenzene	1.781	1.702	4.4	95	0.00
71 M	1,1,2,2-Tetrachloroethane	0.551	0.534	3.1	97	0.00
72	1,2,3-Trichloropropane	0.472	0.431	8.7	93	0.00
73	Bromobenzene	0.431	0.413	4.2	97	0.00
74	n-propylbenzene	2.123	2.073	2.4	98	0.00
75	2-Chlorotoluene	1.319	1.281	2.9	97	0.00
76	1,3,5-Trimethylbenzene	1.500	1.481	1.3	100	0.00
77	t-1,4-Dichloro-2-butene	0.202	0.187	7.4	100	0.00
78	4-Chlorotoluene	1.341	1.299	3.1	98	0.00
79	tert-butylbenzene	1.279	1.245	2.7	99	0.00
80	1,2,4-Trimethylbenzene	1.522	1.452	4.6	97	0.00
81	sec-Butylbenzene	1.762	1.726	2.0	99	0.00
82	p-Isopropyltoluene	1.483	1.429	3.6	99	0.00
83 M	1,3-Dichlorobenzene	0.802	0.774	3.5	98	0.00
84 M	1,4-Dichlorobenzene	0.796	0.772	3.0	102	0.00
85	n-Butylbenzene	1.284	1.218	5.1	102	0.00
86 T	Hexachloroethane	0.286	0.267	6.6	96	0.00
87 M	1,2-Dichlorobenzene	0.775	0.774	0.1	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.113	0.109	3.5	100	0.00
89	1,2,4-Trichlorobenzene	0.403	0.370	8.2	97	0.00
90	Hexachlorobutadiene	0.187	0.184	1.6	106	0.00
91 M	Naphthalene	1.366	1.162	14.9	91	0.00
92	1,2,3-Trichlorobenzene	0.403	0.370	8.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0