

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN100824

Review By	John Carlone	Review On	10/9/2024 9:46:14 AM
Supervise By	Maresh Dadoda	Supervise On	10/9/2024 8:56:13 PM
SubDirectory	VN100824	HP Acquire Method	HP Processing Method 624N100824W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130714		
Initial Calibration Stds	VP130724,VP130725,VP130726,VP130727,VP130728		
CCC	VP130715		
Internal Standard/PEM			
ICV/I.BLK	VP130729		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084336.D	08 Oct 2024 09:08		JC\MD	Ok
2	VSTDICC150	VSTDICC150	VN084337.D	08 Oct 2024 09:43	V13516	JC\MD	Ok,M
3	VSTDICC100	VSTDICC100	VN084338.D	08 Oct 2024 10:07		JC\MD	Ok,M
4	VSTDICC050	VSTDICC050	VN084339.D	08 Oct 2024 10:31		JC\MD	Ok,M
5	VSTDICCC020	VSTDICCC020	VN084340.D	08 Oct 2024 10:55		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN084341.D	08 Oct 2024 11:19		JC\MD	Ok,M
7	IBLK	IBLK	VN084342.D	08 Oct 2024 11:43		JC\MD	Ok
8	VSTDICV020	ICVVN100824	VN084343.D	08 Oct 2024 12:09		JC\MD	Ok,M
9	VN1008WBS01	VN1008WBS01	VN084344.D	08 Oct 2024 12:46		JC\MD	Ok,M
10	VN1008WBSD01	VN1008WBSD01	VN084345.D	08 Oct 2024 13:21		JC\MD	Ok,M
11	VN1008WBL01	VN1008WBL01	VN084346.D	08 Oct 2024 13:45		JC\MD	Ok
12	P4271-01	14B-1	VN084347.D	08 Oct 2024 14:11	vial B pH#6.0	JC\MD	Ok
13	P4334-01	001-WILLETS-PT-BLVD	VN084348.D	08 Oct 2024 14:53	vial A pH<2	JC\MD	Ok,M
14	P4334-02	002-35TH-AVE(SEP)	VN084349.D	08 Oct 2024 15:17	vial A pH<2	JC\MD	Ok
15	P4337-01	001-WILLETS-PT-BLVD	VN084350.D	08 Oct 2024 15:41	vial A pH<2	JC\MD	Ok
16	P4337-02	002-35TH-AVE(OCT)	VN084351.D	08 Oct 2024 16:05	vial A pH<2	JC\MD	Ok
17	P4289-01	EFF-Waste Water	VN084352.D	08 Oct 2024 16:29	vial B pH<2	JC\MD	Ok,M
18	IBLK	IBLK	VN084353.D	08 Oct 2024 16:53		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN100824

Review By	John Carlone	Review On	10/9/2024 9:46:14 AM
Supervise By	Mahesh Dadoda	Supervise On	10/9/2024 8:56:13 PM
SubDirectory	VN100824	HP Acquire Method	HP Processing Method 624N100824W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130714		
Initial Calibration Stds	VP130724,VP130725,VP130726,VP130727,VP130728		
CCC	VP130715		
Internal Standard/PEM			
ICV/I.BLK	VP130729		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	IBLK	IBLK	VN084354.D	08 Oct 2024 17:16		JC\MD	Ok,M
20	VSTDCCC020	VSTDCCC020EC	VN084355.D	08 Oct 2024 17:40		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102524

Review By	Maresh Dadoda	Review On	10/28/2024 2:07:54 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/28/2024 2:09:34 PM
SubDirectory	VN102524	HP Acquire Method	MSVOA_N HP Processing Method 82n093024w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131110,VP131122
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131111,VP131112,VP131123,VP131124 VP128298

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084502.D	25 Oct 2024 08:21		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084503.D	25 Oct 2024 08:59	CCC Failed for comp. #02	JC\MD	Ok,M
3	VSTDCCC020	VSTDCCC020	VN084504.D	25 Oct 2024 10:04		JC\MD	Ok,M
4	VN1025WBS01	VN1025WBS01	VN084505.D	25 Oct 2024 10:40		JC\MD	Ok,M
5	VN1025WBSD01	VN1025WBSD01	VN084506.D	25 Oct 2024 11:04		JC\MD	Ok,M
6	VN1025WBL01	VN1025WBL01	VN084507.D	25 Oct 2024 11:27		JC\MD	Ok
7	VN1025WBL02	VN1025WBL02	VN084508.D	25 Oct 2024 11:51	8260	JC\MD	Ok
8	VN1025WBS02	VN1025WBS02	VN084509.D	25 Oct 2024 12:27	8260	JC\MD	Ok,M
9	VN1025WBSD02	VN1025WBSD02	VN084510.D	25 Oct 2024 12:51	8260	JC\MD	Ok,M
10	P4528-02	241023062-05-TRIP-BL	VN084511.D	25 Oct 2024 13:15	pH#5.0 B TB	JC\MD	Ok,M
11	P4549-01	TT-TB-20241024	VN084512.D	25 Oct 2024 13:39	pH#<2 A TB	JC\MD	Ok
12	P4528-01	241023062-02-VOA	VN084513.D	25 Oct 2024 14:03	pH#7.0 B	JC\MD	Ok
13	VIBLK	VIBLK	VN084514.D	25 Oct 2024 14:27		JC\MD	Ok
14	P4549-03	TT-068-IDWGW-20241	VN084515.D	25 Oct 2024 14:51	bad matrix no straight run	JC\MD	Ok
15	P4549-02	TT-067-IDWGW-20241	VN084516.D	25 Oct 2024 15:15	bad matrix no straight run	JC\MD	Ok
16	P4549-04	TT-069-IDWGW-20241	VN084517.D	25 Oct 2024 15:39	bad matrix no straight run	JC\MD	Ok
17	P4554-01	001-WILLETS-BLVD(SI	VN084518.D	25 Oct 2024 16:03	pH#2.0 A	JC\MD	Ok,M
18	P4550-01	BP-VPB-190-TB-20241	VN084519.D	25 Oct 2024 16:27	pH#2.0 A TB;Surrogate fail	JC\MD	ReRun

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102524

Review By	Mahesh Dadoda	Review On	10/28/2024 2:07:54 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/28/2024 2:09:34 PM		
SubDirectory	VN102524	HP Acquire Method	MSVOA_N	HP Processing Method	82n093024w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131110,VP131122			
CCC		VP131111,VP131112,VP131123,VP131124			
Internal Standard/PEM		VP128298			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4528-02	241023062-05-TRIP-BL	VN084520.D	25 Oct 2024 16:50	pH#5.0 A TB;CCC Failed for comp. #02,no sample to rerun confirmation.no hit on the sample.	JC\MD	Ok
20	P4528-01	241023062-02-VOA	VN084521.D	25 Oct 2024 17:14	pH#7.0 A CCC Failed for comp. #02	JC\MD	ReRun
21	P4550-02	VPB190-HYD-2024102	VN084522.D	25 Oct 2024 17:38	pH#2.0 A Surrogate fail	JC\MD	ReRun
22	P4550-03	BP-VPB-190-EB-20241	VN084523.D	25 Oct 2024 18:02	pH#2.0 A EB	JC\MD	Ok
23	P4550-04	BP-VPB-190-GW-298-3	VN084524.D	25 Oct 2024 18:26	pH#2.0 A	JC\MD	Ok
24	P4550-05	BP-VPB-190-GW-318-3	VN084525.D	25 Oct 2024 18:49	pH#2.0 A Surrogate fail	JC\MD	ReRun
25	VSTDCCC050	VSTDCCC050EC	VN084526.D	25 Oct 2024 19:13		JC\MD	Ok,M
26	VSTDCCC020	VSTDCCC020EC	VN084527.D	25 Oct 2024 19:37		JC\MD	Ok,M

M : Manual Integration