

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4549 SAS No.: P4549 SDG NO.: P4549
Lab File ID: VN084504.D Date Analyzed: 10/25/2024
Instrument ID: MSVOA_N Time Analyzed: 10:04
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	35506	7.81	200200	9.10	180518	11.87
UPPER LIMIT	71012	8.312	400400	9.6	361036	12.365
LOWER LIMIT	17753	7.312	100100	8.6	90259	11.365
EPA SAMPLE NO.						
TT-TB-20241024	29940	7.82	152700	9.10	136002	11.87
TT-067-IDWGW-20241024	28275	7.81	137388	9.10	125598	11.87
TT-068-IDWGW-20241024	29141	7.82	145814	9.10	133742	11.87
TT-069-IDWGW-20241024	27891	7.81	142505	9.10	128076	11.87
VN1025WBL01	32302	7.81	175246	9.10	150285	11.87
VN1025WBS01	35006	7.81	196424	9.10	178483	11.87
VN1025WBSD01	34702	7.81	195213	9.10	177752	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.