

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4550 SAS No.: P4550 SDG No.: P4550

Instrument ID: BNA_N Calibration Date/Time: 10/26/2024 10:16

Lab File ID: BN034734.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.496		-9.8	20.0
Fluoranthene-d10	0.901	0.780		-13.4	20.0
2-Fluorophenol	1.189	1.207		1.5	20.0
Phenol-d6	1.568	1.607		2.5	20.0
Nitrobenzene-d5	0.336	0.310		-7.7	20.0
2-Fluorobiphenyl	1.590	1.390		-12.6	20.0
2,4,6-Tribromophenol	0.124	0.123		-0.8	20.0
Terphenyl-d14	0.817	0.806		-1.3	20.0
1,4-Dioxane	0.532	0.469		-11.8	20.0

All other compounds must meet a minimum RRF of 0.010.