

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4550 SAS No.: P4550 SDG NO.: P4550

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 10/26/2024

Lab File ID: BN034734.D Time Analyzed: 10:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8464	7.647	25235	10.42	12818	14.27
UPPER LIMIT	16928	8.147	50470	10.915	25636	14.772
LOWER LIMIT	4232	7.147	12617.5	9.915	6409	13.772
EPA SAMPLE NO.						
01 BP-VPB-190-GW-398-400	8583	7.65	24972	10.42	11895	14.28
02 PB164409BL	8695	7.65	24402	10.42	10618	14.27
03 VPB190-HYD-20241023RE	9064	7.65	26031	10.42	12214	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.