

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4562 SAS No.: P4562 SDG NO.: P4562

Continuing Calib Date: 10/29/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 13:35 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.43	4.42	4.35	4.49	-0.01
Aroclor-1016-2	(2)	4.45	4.44	4.37	4.51	-0.01
Aroclor-1016-3	(3)	4.50	4.50	4.43	4.57	0.00
Aroclor-1016-4	(4)	4.60	4.59	4.52	4.66	0.00
Aroclor-1016-5	(5)	4.88	4.87	4.80	4.94	-0.01
Aroclor-1260-1	(1)	5.98	5.97	5.90	6.04	-0.01
Aroclor-1260-2	(2)	6.24	6.23	6.16	6.30	-0.01
Aroclor-1260-3	(3)	6.59	6.58	6.51	6.65	-0.01
Aroclor-1260-4	(4)	6.81	6.79	6.72	6.86	-0.01
Aroclor-1260-5	(5)	7.12	7.11	7.04	7.18	-0.01
TCX		3.34	3.33	3.28	3.38	-0.01
DCB		8.47	8.46	8.36	8.56	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4562 SAS No.: P4562 SDG NO.: P4562

Continuing Calib Date: 10/29/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 13:35 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.73	3.74	3.67	3.81	0.01
Aroclor-1016-2	(2)	3.75	3.75	3.68	3.82	0.00
Aroclor-1016-3	(3)	3.91	3.91	3.84	3.98	0.01
Aroclor-1016-4	(4)	3.95	3.96	3.89	4.03	0.01
Aroclor-1016-5	(5)	4.15	4.15	4.08	4.22	0.00
Aroclor-1260-1	(1)	5.14	5.14	5.07	5.21	0.00
Aroclor-1260-2	(2)	5.33	5.33	5.26	5.40	0.01
Aroclor-1260-3	(3)	5.46	5.47	5.40	5.54	0.01
Aroclor-1260-4	(4)	5.92	5.93	5.86	6.00	0.01
Aroclor-1260-5	(5)	6.17	6.18	6.11	6.25	0.01
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.49	7.39	7.59	0.01