

Surrogate Summary

SDG No.: P4562

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4562-08	PCB-GPC-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
P4562-09	PCB-GPC-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
P4562-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
P4562-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150