

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P4562 SAS No.: P4562 SDG NO.: P4562
 Lab File ID: VX043610.D Date Analyzed: 10/30/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	150073	5.54	269353	6.76	239151	10.06
UPPER LIMIT	300146	6.044	538706	7.257	478302	10.555
LOWER LIMIT	75036.5	5.044	134677	6.257	119576	9.555
EPA SAMPLE NO.						
Storage-Blank-WATER-REF-4	140231	5.55	256051	6.76	230067	10.06
Storage-Blank-SAMPLE-MANAGEMENT	140361	5.54	260767	6.76	231351	10.06
VX1030WBL01	124305	5.54	229429	6.76	204357	10.06
VX1030WBS01	148306	5.54	264784	6.76	239021	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.