

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133462

Review By	Sarabjit Jaswal	Review On	11/15/2024 3:07:54 AM
Supervise By	Mohan Bera	Supervise On	11/15/2024 3:09:33 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83050,MP83049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	11/14/24 12:21		Jaswal	OK
2	S00	S0	CAL1	11/14/24 13:10		Jaswal	OK
3	S02	S02	CAL3	11/14/24 13:17		Jaswal	OK
4	S03	S03	CAL4	11/14/24 13:20		Jaswal	OK
5	S04	S04	CAL5	11/14/24 13:23		Jaswal	OK
6	S05	S05	CAL6	11/14/24 13:26		Jaswal	OK
7	S06	S06	CAL7	11/14/24 13:29		Jaswal	OK
8	S07	S07	CAL8	11/14/24 13:31		Jaswal	OK
9	S08	S08	CAL9	11/14/24 13:34		Jaswal	OK
10	ICV001	ICV001	ICV	11/14/24 15:53	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	11/14/24 16:04	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	11/14/24 16:19	Inst. paused for review	Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	11/14/24 16:35	Inst. paused for review	Jaswal	OK
14	CCV001	CCV001	CCV	11/14/24 16:38		Jaswal	OK
15	CCB001	CCB001	CCB	11/14/24 16:41		Jaswal	OK
16	P4570-01A	MYE531ARE	PS	11/14/24 16:44	PS For Ag- 0.02 - MP83071	Jaswal	OK
17	CCV002	CCV002	CCV	11/14/24 16:57		Jaswal	OK
18	CCB002	CCB002	CCB	11/14/24 17:00		Jaswal	OK