

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE543
 Instrument ID: P7 Date: 11/10/2024
 Run Batch: LB133382

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.8	5	0.41
Co - 59	59.0	0.8	5	1.2
In - 113	113.0	0.7	5	0.82
In - 115	115.0	0.8	5	1.7
Mg - 24	23.9	0.8	5	0.81
Mg - 25	24.9	0.8	5	0.38
Mg - 26	25.9	0.8	5	0.28
Pb - 206	206.1	0.8	5	1.0
Pb - 207	207.0	0.8	5	0.94
Pb - 208	208.1	0.8	5	0.44

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE543
 Instrument ID: P7 Date: 11/12/2024
 Run Batch: LB133416

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.0	0.7	5	0.33
Co - 59	59.0	0.8	5	0.66
In - 113	113.0	0.7	5	0.51
In - 115	115.0	0.8	5	0.47
Mg - 24	24.0	0.8	5	0.69
Mg - 25	25.0	0.8	5	0.39
Mg - 26	26.0	0.8	5	0.33
Pb - 206	206.0	0.8	5	0.59
Pb - 207	207.0	0.8	5	0.57
Pb - 208	208.0	0.8	5	1.5

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE543
 Instrument ID: P7 Date: 11/13/2024
 Run Batch: LB133439

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.0	0.8	5	0.16
Co - 59	58.9	0.8	5	0.97
In - 113	113.0	0.8	5	0.60
In - 115	115.0	0.8	5	0.48
Mg - 24	23.9	0.8	5	0.71
Mg - 25	24.9	0.8	5	0.16
Mg - 26	25.9	0.8	5	0.19
Pb - 206	206.0	0.8	5	0.28
Pb - 207	207.0	0.8	5	0.54
Pb - 208	208.0	0.8	5	0.78