

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True   |        | Found |    |       |    |
|----------|--------|--------|-------|----|-------|----|
|          | ICSA   | ICSAB  | ICSA  | %R | ICSAB | %R |
| Aluminum | 100000 | 100000 | 95000 | 95 | 95000 | 95 |
| Aluminum | 100000 | 100000 | 95000 | 95 | 93000 | 93 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True   |        | Found |    |       |    |
|----------|--------|--------|-------|----|-------|----|
|          | ICSA   | ICSAB  | ICSA  | %R | ICSAB | %R |
| Aluminum | 100000 | 100000 | 88000 | 88 | 95000 | 95 |
| Aluminum | 100000 | 100000 | 88000 | 88 | 93000 | 93 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Antimony | 1.5  | 22    | 1.2   | 79 | 20    | 92 |
| Antimony | 1.5  | 22    | 1.2   | 79 | 19    | 88 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Antimony | 1.5  | 22    | 1.1   | 70 | 20    | 92 |
| Antimony | 1.5  | 22    | 1.1   | 70 | 19    | 88 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Arsenic | 0.1  | 19    | 0.35  | 350 | 19    | 101 |
| Arsenic | 0.1  | 19    | 0.35  | 350 | 20    | 103 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Arsenic | 0.1  | 19    | 0.29  | 290 | 20    | 103 |
| Arsenic | 0.1  | 19    | 0.29  | 290 | 19    | 101 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |    |
|---------|------|-------|-------|-----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Barium  | 1.2  | 22    | 1.5   | 123 | 20    | 90 |
| Barium  | 1.2  | 22    | 1.5   | 123 | 21    | 93 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |    |
|---------|------|-------|-------|-----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Barium  | 1.2  | 22    | 1.4   | 113 | 21    | 93 |
| Barium  | 1.2  | 22    | 1.4   | 113 | 20    | 90 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True |       | Found |    |       |     |
|-----------|------|-------|-------|----|-------|-----|
|           | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Beryllium | 0    | 19    | 0.27  |    | 19    | 100 |
| Beryllium | 0    | 19    | 0.27  |    | 20    | 103 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True |       | Found |    |       |     |
|-----------|------|-------|-------|----|-------|-----|
|           | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Beryllium | 0    | 19    | 0.27  |    | 20    | 103 |
| Beryllium | 0    | 19    | 0.27  |    | 19    | 100 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |    |
|---------|------|-------|-------|----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Cadmium | 0.7  | 20    | 0.41  | 59 | 18    | 88 |
| Cadmium | 0.7  | 20    | 0.41  | 59 | 19    | 95 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |    |
|---------|------|-------|-------|----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Cadmium | 0.7  | 20    | 0.33  | 47 | 19    | 95 |
| Cadmium | 0.7  | 20    | 0.33  | 47 | 18    | 88 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found |    |       |    |
|---------|--------|--------|-------|----|-------|----|
|         | ICSA   | ICSAB  | ICSA  | %R | ICSAB | %R |
| Calcium | 100000 | 100000 | 88000 | 88 | 89000 | 89 |
| Calcium | 100000 | 100000 | 88000 | 88 | 88000 | 88 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found |    |       |    |
|---------|--------|--------|-------|----|-------|----|
|         | ICSA   | ICSAB  | ICSA  | %R | ICSAB | %R |
| Calcium | 100000 | 100000 | 83000 | 83 | 88000 | 88 |
| Calcium | 100000 | 100000 | 83000 | 83 | 89000 | 89 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Chromium | 21   | 40    | 21    | 98 | 39    | 97 |
| Chromium | 21   | 40    | 21    | 98 | 40    | 99 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Chromium | 21   | 40    | 20    | 94 | 39    | 97 |
| Chromium | 21   | 40    | 20    | 94 | 40    | 99 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Cobalt  | 1    | 20    | 1.7   | 174 | 20    | 101 |
| Cobalt  | 1    | 20    | 1.7   | 174 | 21    | 104 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Cobalt  | 1    | 20    | 1.7   | 168 | 20    | 101 |
| Cobalt  | 1    | 20    | 1.7   | 168 | 21    | 104 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Copper  | 8    | 25    | 7.6   | 94 | 25    | 101 |
| Copper  | 8    | 25    | 7.6   | 94 | 25    | 98  |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Copper  | 8    | 25    | 7.1   | 89 | 25    | 98  |
| Copper  | 8    | 25    | 7.1   | 89 | 25    | 101 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found  |     |        |     |
|---------|--------|--------|--------|-----|--------|-----|
|         | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Iron    | 100000 | 100000 | 110000 | 105 | 110000 | 106 |
| Iron    | 100000 | 100000 | 110000 | 105 | 110000 | 105 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found  |     |        |     |
|---------|--------|--------|--------|-----|--------|-----|
|         | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Iron    | 100000 | 100000 | 100000 | 100 | 110000 | 105 |
| Iron    | 100000 | 100000 | 100000 | 100 | 110000 | 106 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |    |
|---------|------|-------|-------|-----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Lead    | 4    | 25    | 4.8   | 120 | 23    | 93 |
| Lead    | 4    | 25    | 4.8   | 120 | 23    | 93 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |    |
|---------|------|-------|-------|-----|-------|----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Lead    | 4    | 25    | 4.2   | 105 | 23    | 93 |
| Lead    | 4    | 25    | 4.2   | 105 | 23    | 93 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Magnesium | 100000 | 100000 | 100000 | 103 | 100000 | 104 |
| Magnesium | 100000 | 100000 | 100000 | 103 | 100000 | 104 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True   |        | Found |    |        |     |
|-----------|--------|--------|-------|----|--------|-----|
|           | ICSA   | ICSAB  | ICSA  | %R | ICSAB  | %R  |
| Magnesium | 100000 | 100000 | 99000 | 99 | 100000 | 104 |
| Magnesium | 100000 | 100000 | 99000 | 99 | 100000 | 104 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True |       | Found |     |       |    |
|-----------|------|-------|-------|-----|-------|----|
|           | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Manganese | 7    | 27    | 8.0   | 114 | 26    | 97 |
| Manganese | 7    | 27    | 8.0   | 114 | 26    | 97 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True |       | Found |     |       |    |
|-----------|------|-------|-------|-----|-------|----|
|           | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R |
| Manganese | 7    | 27    | 7.5   | 107 | 26    | 97 |
| Manganese | 7    | 27    | 7.5   | 107 | 26    | 97 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Nickel  | 6    | 24    | 5.5   | 91 | 24    | 98  |
| Nickel  | 6    | 24    | 5.5   | 91 | 24    | 101 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Nickel  | 6    | 24    | 5.3   | 89 | 24    | 98  |
| Nickel  | 6    | 24    | 5.3   | 89 | 24    | 101 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Potassium | 100000 | 100000 | 100000 | 101 | 98000  | 98  |
| Potassium | 100000 | 100000 | 100000 | 101 | 100000 | 101 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte   | True   |        | Found |    |        |     |
|-----------|--------|--------|-------|----|--------|-----|
|           | ICSA   | ICSAB  | ICSA  | %R | ICSAB  | %R  |
| Potassium | 100000 | 100000 | 92000 | 92 | 100000 | 101 |
| Potassium | 100000 | 100000 | 92000 | 92 | 98000  | 98  |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Selenium | 0.3  | 19    | 0     | 0  | 18    | 96 |
| Selenium | 0.3  | 19    | 0     | 0  | 18    | 96 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Selenium | 0.3  | 19    | -0.4  | 0  | 18    | 96 |
| Selenium | 0.3  | 19    | -0.4  | 0  | 18    | 96 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Silver  | 0    | 18    | 0.02  |    | 17    | 96  |
| Silver  | 0    | 18    | 0.02  |    | 18    | 102 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |    |       |     |
|---------|------|-------|-------|----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R | ICSAB | %R  |
| Silver  | 0    | 18    | 0.01  |    | 18    | 102 |
| Silver  | 0    | 18    | 0.01  |    | 17    | 96  |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found  |     |        |     |
|---------|--------|--------|--------|-----|--------|-----|
|         | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Sodium  | 100000 | 100000 | 100000 | 102 | 100000 | 103 |
| Sodium  | 100000 | 100000 | 100000 | 102 | 100000 | 102 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True   |        | Found |    |        |     |
|---------|--------|--------|-------|----|--------|-----|
|         | ICSA   | ICSAB  | ICSA  | %R | ICSAB  | %R  |
| Sodium  | 100000 | 100000 | 97000 | 97 | 100000 | 102 |
| Sodium  | 100000 | 100000 | 97000 | 97 | 100000 | 103 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Thallium | 0    | 21    | 0.15  |    | 19    | 92 |
| Thallium | 0    | 21    | 0.15  |    | 19    | 90 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Thallium | 0    | 21    | 0.01  |    | 19    | 90 |
| Thallium | 0    | 21    | 0.01  |    | 19    | 92 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Vanadium | 0.5  | 19    | 0.19  | 38 | 19    | 98 |
| Vanadium | 0.5  | 19    | 0.19  | 38 | 19    | 99 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte  | True |       | Found |    |       |    |
|----------|------|-------|-------|----|-------|----|
|          | ICSA | ICSAB | ICSA  | %R | ICSAB | %R |
| Vanadium | 0.5  | 19    | 0.19  | 38 | 19    | 98 |
| Vanadium | 0.5  | 19    | 0.19  | 38 | 19    | 99 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Zinc    | 11   | 29    | 11    | 103 | 30    | 103 |
| Zinc    | 11   | 29    | 11    | 103 | 31    | 106 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
Analytical Method: ICP-MS ICSA Source: EPA-0803  
Instrument ID: P7 ICSB Source: EPA-0803  
Run Batch: LB133382  
Concentration Units: ug/L

| Analyte | True |       | Found |     |       |     |
|---------|------|-------|-------|-----|-------|-----|
|         | ICSA | ICSAB | ICSA  | %R  | ICSAB | %R  |
| Zinc    | 11   | 29    | 12    | 105 | 30    | 103 |
| Zinc    | 11   | 29    | 12    | 105 | 31    | 106 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
 Analytical Method: ICP-MS ICSA Source: EPA-0803  
 Instrument ID: P7 ICSB Source: EPA-0803  
 Run Batch: LB133416  
 Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Aluminum  | 100000 | 100000 | 93000  | 93  | 90000  | 90  |
| Antimony  | 1.5    | 22     | 1.1    | 76  | 21     | 97  |
| Arsenic   | 0.1    | 19     | 0.32   | 320 | 20     | 106 |
| Barium    | 1.2    | 22     | 1.5    | 122 | 21     | 97  |
| Beryllium | 0      | 19     | 0.3    |     | 20     | 107 |
| Cadmium   | 0.7    | 20     | 0.39   | 56  | 21     | 102 |
| Calcium   | 100000 | 100000 | 100000 | 102 | 100000 | 100 |
| Chromium  | 21     | 40     | 20     | 95  | 39     | 98  |
| Cobalt    | 1      | 20     | 1.3    | 125 | 21     | 104 |
| Copper    | 8      | 25     | 7.1    | 89  | 26     | 102 |
| Iron      | 100000 | 100000 | 100000 | 104 | 100000 | 103 |
| Lead      | 4      | 25     | 4.6    | 115 | 25     | 98  |
| Magnesium | 100000 | 100000 | 100000 | 101 | 98000  | 98  |
| Manganese | 7      | 27     | 8.0    | 114 | 27     | 101 |
| Nickel    | 6      | 24     | 5.3    | 88  | 25     | 103 |
| Potassium | 100000 | 100000 | 100000 | 101 | 99000  | 99  |
| Selenium  | 0.3    | 19     | 0.4    | 133 | 19     | 101 |
| Silver    | 0      | 18     | 0.03   |     | 19     | 107 |
| Sodium    | 100000 | 100000 | 97000  | 97  | 96000  | 96  |
| Thallium  | 0      | 21     | 0.12   |     | 21     | 98  |
| Vanadium  | 0.5    | 19     | 0.17   | 34  | 19     | 101 |
| Zinc      | 11     | 29     | 11     | 99  | 30     | 105 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51817 MA No. :                      SDG No.: MYE543  
 Analytical Method: ICP-MS ICSA Source: EPA-0803  
 Instrument ID: P7 ICSB Source: EPA-0803  
 Run Batch: LB133439  
 Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Aluminum  | 100000 | 100000 | 92000  | 92  | 92000  | 92  |
| Antimony  | 1.5    | 22     | 1.1    | 74  | 21     | 97  |
| Arsenic   | 0.1    | 19     | 0.34   | 340 | 21     | 109 |
| Barium    | 1.2    | 22     | 1.4    | 118 | 22     | 98  |
| Beryllium | 0      | 19     | 0.28   |     | 21     | 112 |
| Cadmium   | 0.7    | 20     | 0.11   | 16  | 20     | 101 |
| Calcium   | 100000 | 100000 | 100000 | 102 | 100000 | 104 |
| Chromium  | 21     | 40     | 21     | 97  | 41     | 102 |
| Cobalt    | 1      | 20     | 1.3    | 127 | 22     | 109 |
| Copper    | 8      | 25     | 7.4    | 93  | 27     | 106 |
| Iron      | 100000 | 100000 | 100000 | 104 | 100000 | 104 |
| Lead      | 4      | 25     | 4.5    | 112 | 25     | 99  |
| Magnesium | 100000 | 100000 | 100000 | 100 | 99000  | 99  |
| Manganese | 7      | 27     | 8.1    | 116 | 28     | 104 |
| Nickel    | 6      | 24     | 5.5    | 91  | 26     | 107 |
| Potassium | 100000 | 100000 | 99000  | 99  | 99000  | 99  |
| Selenium  | 0.3    | 19     | 0      | 0   | 20     | 106 |
| Silver    | 0      | 18     | 0.02   |     | 19     | 108 |
| Sodium    | 100000 | 100000 | 100000 | 100 | 100000 | 99  |
| Thallium  | 0      | 21     | 0.08   |     | 21     | 98  |
| Vanadium  | 0.5    | 19     | 0.18   | 36  | 20     | 106 |
| Zinc      | 11     | 29     | 12     | 107 | 33     | 112 |