

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	11/12/24 11:12		Jaswal	OK
2	S00	S0	CAL1	11/12/24 12:38		Jaswal	OK
3	S02	S02	CAL3	11/12/24 12:45		Jaswal	OK
4	S03	S03	CAL4	11/12/24 12:48		Jaswal	OK
5	S04	S04	CAL5	11/12/24 12:51		Jaswal	OK
6	S05	S05	CAL6	11/12/24 12:54		Jaswal	OK
7	S06	S06	CAL7	11/12/24 12:56		Jaswal	OK
8	S07	S07	CAL8	11/12/24 12:59		Jaswal	OK
9	S08	S08	CAL9	11/12/24 13:02		Jaswal	OK
10	ICV001	ICV001	ICV	11/12/24 13:43	Inst. paused Calibration for review	Jaswal	OK
11	ICB001	ICB001	ICB	11/12/24 13:46		Jaswal	OK
12	ICSA001	ICSA001	ICSA	11/12/24 13:49		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	11/12/24 13:52		Jaswal	OK
14	CCV001	CCV001	CCV	11/12/24 13:59	Inst. paused for review	Jaswal	OK
15	CCB001	CCB001	CCB	11/12/24 14:02		Jaswal	OK
16	PB164655BL	PBS655	MB	11/12/24 14:07		Jaswal	OK
17	PB164655BS	LCS655	LCS	11/12/24 14:10		Jaswal	OK
18	PB164655BSD	LCSD655	LCSD	11/12/24 14:13		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

19	PB164657BL	PBS657	MB	11/12/24 14:17		Jaswal	OK
20	PB164657BS	LCS657	LCS	11/12/24 14:20		Jaswal	OK
21	PB164657BSD	LCSD657	LCSD	11/12/24 14:23		Jaswal	OK
22	PB164659BL	PBS659	MB	11/12/24 14:26		Jaswal	OK
23	PB164659BS	LCS659	LCS	11/12/24 14:30		Jaswal	OK
24	PB164659BSD	LCSD659	LCSD	11/12/24 14:33		Jaswal	OK
25	P4572-01	MYE557	SAM	11/12/24 14:36		Jaswal	OK
26	P4572-02	MYE558	SAM	11/12/24 14:39		Jaswal	OK
27	P4572-03	MYE559	SAM	11/12/24 14:42		Jaswal	OK
28	P4572-04	MYE560	SAM	11/12/24 14:45		Jaswal	OK
29	P4572-05	MYE561	SAM	11/12/24 14:48		Jaswal	OK
30	P4572-06	MYE562	SAM	11/12/24 14:51		Jaswal	OK
31	P4572-07	MYE582	SAM	11/12/24 14:54		Jaswal	OK
32	P4572-08	MYE583	SAM	11/12/24 14:57		Jaswal	OK
33	P4572-09	MYE584	SAM	11/12/24 15:00		Jaswal	OK
34	P4572-10	MYE585	SAM	11/12/24 15:04		Jaswal	OK
35	P4572-11	MYE586	SAM	11/12/24 15:07	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
36	P4572-12	MYE586D	DUP	11/12/24 15:10	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
37	P4572-11L	MYE586L	SD	11/12/24 15:13	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
38	P4572-13	MYE586S	MS	11/12/24 15:16	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

39	P4572-11MS	MYE586SRE	MS	11/12/24 15:19	Ba,Co,Ni,V High	Jaswal	Dilution
40	P4572-14	MYE587	SAM	11/12/24 15:21		Jaswal	OK
41	P4572-15	MYE588	SAM	11/12/24 15:24		Jaswal	OK
42	P4572-16	MYE589	SAM	11/12/24 15:28		Jaswal	OK
43	P4572-17	MYE590	SAM	11/12/24 15:30		Jaswal	OK
44	P4572-18	MYE591	SAM	11/12/24 15:33		Jaswal	OK
45	P4572-19	MYE592	SAM	11/12/24 15:36		Jaswal	OK
46	P4572-20	MYE593	SAM	11/12/24 15:39		Jaswal	OK
47	CCV002	CCV002	CCV	11/12/24 15:42		Jaswal	OK
48	CCB002	CCB002	CCB	11/12/24 15:45		Jaswal	OK
49	P4572-21	MYE594	SAM	11/12/24 15:48		Jaswal	OK
50	P4572-22	MYE595	SAM	11/12/24 15:51		Jaswal	OK
51	P4573-01	MYE596	SAM	11/12/24 15:54		Jaswal	OK
52	P4573-02	MYE597	SAM	11/12/24 15:57		Jaswal	OK
53	P4573-03	MYE598	SAM	11/12/24 16:00		Jaswal	OK
54	P4573-04	MYE599	SAM	11/12/24 16:03		Jaswal	OK
55	P4573-05	MYE5A0	SAM	11/12/24 16:06		Jaswal	OK
56	P4573-06	MYE5A1	SAM	11/12/24 16:10		Jaswal	OK
57	P4573-07	MYE5A2	SAM	11/12/24 16:13	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
58	P4573-08	MYE5A2D	DUP	11/12/24 16:16	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

59	P4573-07L	MYE5A2L	SD	11/12/24 16:19	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
60	P4573-09	MYE5A2S	MS	11/12/24 16:22	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
61	P4573-07MS	MYE5A2SRE	MS	11/12/24 16:25	MS FailI for Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
62	P4573-10	MYE5A3	SAM	11/12/24 16:28		Jaswal	OK
63	P4573-11	MYE5A4	SAM	11/12/24 16:31		Jaswal	OK
64	P4573-12	MYE5A5	SAM	11/12/24 16:34		Jaswal	OK
65	P4573-13	MYE5A6	SAM	11/12/24 16:37		Jaswal	OK
66	P4573-14	MYE5A7	SAM	11/12/24 16:40		Jaswal	OK
67	P4573-15	MYE5A8	SAM	11/12/24 16:43		Jaswal	OK
68	P4573-16	MYE5A9	SAM	11/12/24 16:46		Jaswal	OK
69	P4573-17	MYE5B0	SAM	11/12/24 16:49		Jaswal	OK
70	P4573-18	MYE5B1	SAM	11/12/24 16:52		Jaswal	OK
71	P4573-19	MYE5B2	SAM	11/12/24 16:55		Jaswal	OK
72	P4587-01	MYE5G4	SAM	11/12/24 16:58		Jaswal	OK
73	P4587-02	MYE5G5	SAM	11/12/24 17:01		Jaswal	OK
74	P4587-03	MYE5G6	SAM	11/12/24 17:04		Jaswal	OK
75	CCV003	CCV003	CCV	11/12/24 17:07		Jaswal	OK
76	CCB003	CCB003	CCB	11/12/24 17:10		Jaswal	OK
77	P4587-04	MYE5G7	SAM	11/12/24 17:13		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

78	P4587-05	MYE5G8	SAM	11/12/24 17:16		Jaswal	OK
79	P4587-06	MYE5G9	SAM	11/12/24 17:19	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
80	P4587-07	MYE5G9D	DUP	11/12/24 17:22	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
81	P4587-06L	MYE5G9L	SD	11/12/24 17:25	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
82	P4587-08	MYE5G9S	MS	11/12/24 17:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
83	P4587-06MS	MYE5G9SRE	MS	11/12/24 17:31	Ba,Co,Ni,V High	Jaswal	Dilution
84	P4587-09	MYE5H0	SAM	11/12/24 17:34		Jaswal	OK
85	P4587-10	MYE5H1	SAM	11/12/24 17:37		Jaswal	OK
86	P4587-11	MYE5H2	SAM	11/12/24 17:40		Jaswal	OK
87	P4587-12	MYE5H3	SAM	11/12/24 17:43		Jaswal	OK
88	P4587-13	MYE5H4	SAM	11/12/24 17:46		Jaswal	OK
89	P4587-14	MYE5H5	SAM	11/12/24 17:49		Jaswal	OK
90	P4587-15	MYE5H6	SAM	11/12/24 17:52		Jaswal	OK
91	P4587-16	MYE5H7	SAM	11/12/24 17:55		Jaswal	OK
92	P4587-17	MYE5H8	SAM	11/12/24 17:58		Jaswal	OK
93	P4587-18	MYE5H9	SAM	11/12/24 18:01		Jaswal	OK
94	P4587-19	MYE5J0	SAM	11/12/24 18:05		Jaswal	OK
95	P4587-20	MYE5J1	SAM	11/12/24 18:08		Jaswal	OK
96	P4572-11DL	MYE586	SAM	11/12/24 18:11	4X for Ba,Co,Ni,V	Jaswal	Confirms
97	P4572-12DL	MYE586D	DUP	11/12/24 18:14	4X for Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

98	P4572-11LDL	MYE586L	SD	11/12/24 18:17	20X for Ba,Co,Ni,V	Jaswal	Confirms
99	P4572-13DL	MYE586S	MS	11/12/24 18:20	4X for Ba,Co,Ni,V	Jaswal	Confirms
100	P4572-11MSDL	MYE586SRE	MS	11/12/24 18:23	4X for Ba,Co,Ni,V	Jaswal	Confirms
101	P4573-07DL	MYE5A2	SAM	11/12/24 18:26	4X for Ba,Co,Ni,V	Jaswal	Confirms
102	P4573-08DL	MYE5A2D	DUP	11/12/24 18:29	4X for Ba,Co,Ni,V	Jaswal	Confirms
103	P4573-07LDL	MYE5A2L	SD	11/12/24 18:33	20X for Ba,Co,Ni,V	Jaswal	Confirms
104	P4573-09DL	MYE5A2S	MS	11/12/24 18:36	4X for Ba,Co,Ni,V	Jaswal	Confirms
105	CCV004	CCV004	CCV	11/12/24 18:39		Jaswal	OK
106	CCB004	CCB004	CCB	11/12/24 18:42		Jaswal	OK
107	P4573-07MSDL	MYE5A2SRE	MS	11/12/24 18:45	4X for Ba,Co,Ni,V	Jaswal	Confirms
108	P4587-06DL	MYE5G9	SAM	11/12/24 18:48	4X for Ba,Co,Ni,V	Jaswal	Confirms
109	P4587-07DL	MYE5G9D	DUP	11/12/24 18:51	4X for Ba,Co,Ni,V	Jaswal	Confirms
110	P4587-06LDL	MYE5G9L	SD	11/12/24 18:54	20X for Ba,Co,Ni,V	Jaswal	Confirms
111	P4587-08DL	MYE5G9S	MS	11/12/24 18:57	4X for Ba,Co,Ni,V	Jaswal	Confirms
112	P4587-06MSDL	MYE5G9SRE	MS	11/12/24 19:00	4X for Ba,Co,Ni,V	Jaswal	Confirms
113	P4571-01DL	MYE543	SAM	11/12/24 19:12	MS Fail for Many Parameter so please confirm	Jaswal	Not Ok
114	P4571-02DL	MYE543D	DUP	11/12/24 19:15	MS Fail for Many Parameter so please confirm	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

115	P4571-01LDL	MYE543L	SD	11/12/24 19:19	MS Fail for Many Parameter so please confirm	Jaswal	Not Ok
116	P4571-03DL	MYE543S	MS	11/12/24 19:22	MS Fail for Many Parameter so please confirm	Jaswal	Not Ok
117	P4571-01MSDL	MYE543SRE	MS	11/12/24 19:25	MS Fail for Many Parameter so please confirm	Jaswal	Not Ok
118	P4571-01A	MYE543ARE	PS	11/12/24 19:28	Not Required	Jaswal	Not Ok
119	P4571-05	MYE577	SAM	11/12/24 19:31		Jaswal	OK
120	P4571-06	MYE578	SAM	11/12/24 19:34		Jaswal	OK
121	P4571-20DL	MYE554	SAM	11/12/24 19:37	Not Use	Jaswal	Not Ok
122	PB164829BL	PBW829	MB	11/12/24 19:41		Jaswal	OK
123	PB164829BS	LCS829	LCS	11/12/24 19:44		Jaswal	OK
124	P4787-01	GCP96	SAM	11/12/24 19:48		Jaswal	OK
125	P4787-02	GCP96D	DUP	11/12/24 19:51		Jaswal	OK
126	P4787-01L	GCP96L	SD	11/12/24 19:54		Jaswal	OK
127	P4787-03	GCP96S	MS	11/12/24 19:57		Jaswal	OK
128	P4787-04	GCP97	SAM	11/12/24 20:00		Jaswal	OK
129	P4787-05	GCP98	SAM	11/12/24 20:03		Jaswal	OK
130	P4787-06	GCP99	SAM	11/12/24 20:07		Jaswal	OK
131	P4787-07	GCPA0	SAM	11/12/24 20:10		Jaswal	OK
132	P4787-08	GCPA1	SAM	11/12/24 20:13	V High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

133	P4787-08DL	GCPA1	SAM	11/12/24 20:16	2X For V	Jaswal	Confirms
134	CCV005	CCV005	CCV	11/12/24 20:19		Jaswal	OK
135	CCB005	CCB005	CCB	11/12/24 20:22		Jaswal	OK
136	P4787-08DL2	GCPA1	SAM	11/12/24 20:25	Not Required	Jaswal	Not Ok
137	P4787-08DL2RE	GCPA1	SAM	11/12/24 20:28	Not Required	Jaswal	Not Ok
138	PB164716BL	PBS716	MB	11/12/24 20:31		Jaswal	OK
139	PB164716BS	LCS716	LCS	11/12/24 20:39		Jaswal	OK
140	PB164716BSD	LCSD716	LCSD	11/12/24 20:42		Jaswal	OK
141	PB164699BL	PBS699	MB	11/12/24 20:45		Jaswal	OK
142	PB164699BS	LCS699	LCS	11/12/24 20:48		Jaswal	OK
143	PB164699BSD	LCSD699	LCSD	11/12/24 20:52		Jaswal	OK
144	P4590-01	MYE5N3	SAM	11/12/24 20:55	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
145	P4590-02	MYE5N3D	DUP	11/12/24 20:58	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
146	P4590-01L	MYE5N3L	SD	11/12/24 21:01	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
147	P4590-03	MYE5N3S	MS	11/12/24 21:04	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
148	P4590-01MS	MYE5N3SRE	MS	11/12/24 21:07	Ba,Co,Ni,V High	Jaswal	Dilution
149	P4591-01	MYE5N4	SAM	11/12/24 21:10	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
150	P4591-02	MYE5N4D	DUP	11/12/24 21:13	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
151	P4591-01L	MYE5N4L	SD	11/12/24 21:16	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
152	P4591-03	MYE5N4S	MS	11/12/24 21:19	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

153	P4591-01MS	MYE5N4SRE	MS	11/12/24 21:22	Ba,Co,Ni,V High	Jaswal	Dilution
154	P4591-04	MYE5N5	SAM	11/12/24 21:25		Jaswal	OK
155	P4589-01	MYE5L0	SAM	11/12/24 21:28		Jaswal	OK
156	P4589-02	MYE5L1	SAM	11/12/24 21:31		Jaswal	OK
157	P4589-03	MYE5L2	SAM	11/12/24 21:34		Jaswal	OK
158	P4589-04	MYE5M3	SAM	11/12/24 21:38		Jaswal	OK
159	P4589-05	MYE5M4	SAM	11/12/24 21:41		Jaswal	OK
160	P4589-06	MYE5M5	SAM	11/12/24 21:44		Jaswal	OK
161	P4589-07	MYE5M6	SAM	11/12/24 21:47		Jaswal	OK
162	P4589-08	MYE5M7	SAM	11/12/24 21:50		Jaswal	OK
163	P4589-09	MYE5M8	SAM	11/12/24 21:53		Jaswal	OK
164	CCV006	CCV006	CCV	11/12/24 21:58		Jaswal	OK
165	CCB006	CCB006	CCB	11/12/24 22:03		Jaswal	OK
166	P4589-10	MYE5M9	SAM	11/12/24 22:06		Jaswal	OK
167	P4589-11	MYE5N0	SAM	11/12/24 22:09		Jaswal	OK
168	P4589-12	MYE5N1	SAM	11/12/24 22:12		Jaswal	OK
169	P4589-13	MYE5N2	SAM	11/12/24 22:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
170	P4589-14	MYE5N2D	DUP	11/12/24 22:18	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
171	P4589-13L	MYE5N2L	SD	11/12/24 22:21	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
172	P4589-15	MYE5N2S	MS	11/12/24 22:24	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

173	P4589-13MS	MYE5N2SRE	MS	11/12/24 22:27	Ba,Co,Ni,V High	Jaswal	Dilution
174	PB164696BL	PBS696	MB	11/12/24 22:30		Jaswal	OK
175	PB164696BS	LCS696	LCS	11/12/24 22:33		Jaswal	OK
176	PB164696BSD	LCSD696	LCSD	11/12/24 22:37		Jaswal	OK
177	P4588-01	MYE5J2	SAM	11/12/24 22:40		Jaswal	OK
178	P4588-02	MYE5J3	SAM	11/12/24 22:43		Jaswal	OK
179	P4588-03	MYE5J4	SAM	11/12/24 22:46		Jaswal	OK
180	P4588-04	MYE5J5	SAM	11/12/24 22:49		Jaswal	OK
181	P4588-05	MYE5J6	SAM	11/12/24 22:52		Jaswal	OK
182	P4588-06	MYE5J7	SAM	11/12/24 22:55		Jaswal	OK
183	P4588-07	MYE5J8	SAM	11/12/24 22:59		Jaswal	OK
184	P4588-08	MYE5J9	SAM	11/12/24 23:02		Jaswal	OK
185	P4588-09	MYE5K0	SAM	11/12/24 23:05		Jaswal	OK
186	P4588-10	MYE5K1	SAM	11/12/24 23:08		Jaswal	OK
187	P4588-11	MYE5K2	SAM	11/12/24 23:11		Jaswal	OK
188	P4588-12	MYE5K3	SAM	11/12/24 23:14		Jaswal	OK
189	P4588-13	MYE5K4	SAM	11/12/24 23:17		Jaswal	OK
190	P4588-14	MYE5K5	SAM	11/12/24 23:20	Not USe	Jaswal	Not Ok
191	P4588-15	MYE5K5D	DUP	11/12/24 23:23	Not USe	Jaswal	Not Ok
192	P4588-14L	MYE5K5L	SD	11/12/24 23:26	Not USe	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

193	P4588-16	MYE5K5S	MS	11/12/24 23:29	Not USe	Jaswal	Not Ok
194	CCV007	CCV007	CCV	11/12/24 23:32		Jaswal	OK
195	CCB007	CCB007	CCB	11/12/24 23:35		Jaswal	OK
196	P4588-14MS	MYE5K5SRE	MS	11/12/24 23:38	Not USe	Jaswal	Not Ok
197	P4588-17	MYE5K6	SAM	11/12/24 23:41		Jaswal	OK
198	P4588-18	MYE5K7	SAM	11/12/24 23:44		Jaswal	OK
199	P4588-19	MYE5K8	SAM	11/12/24 23:47		Jaswal	OK
200	P4588-20	MYE5K9	SAM	11/12/24 23:50		Jaswal	OK
201	P4590-01DL	MYE5N3	SAM	11/12/24 23:53	4X for Ba,Co,Ni,V	Jaswal	Confirms
202	P4590-02DL	MYE5N3D	DUP	11/12/24 23:57	4X for Ba,Co,Ni,V	Jaswal	Confirms
203	P4590-01LDL	MYE5N3L	SD	11/13/24	20X for Ba,Co,Ni,V	Jaswal	Confirms
204	P4590-03DL	MYE5N3S	MS	11/13/24 00:03	4X for Ba,Co,Ni,V	Jaswal	Confirms
205	P4590-01MSDL	MYE5N3SRE	MS	11/13/24 00:06	4X for Ba,Co,Ni,V	Jaswal	Confirms
206	P4591-01DL	MYE5N4	SAM	11/13/24 00:09	4X for Ba,Co,Ni,V	Jaswal	Confirms
207	P4591-02DL	MYE5N4D	DUP	11/13/24 00:12	4X for Ba,Co,Ni,V	Jaswal	Confirms
208	P4591-01LDL	MYE5N4L	SD	11/13/24 00:15	20X for Ba,Co,Ni,V	Jaswal	Confirms
209	P4591-03DL	MYE5N4S	MS	11/13/24 00:19	4X for Ba,Co,Ni,V	Jaswal	Confirms
210	P4591-01MSDL	MYE5N4SRE	MS	11/13/24 00:22	4X for Ba,Co,Ni,V	Jaswal	Confirms
211	P4589-13DL	MYE5N2	SAM	11/13/24 00:25	4X for Ba,Co,Ni,V	Jaswal	Confirms
212	P4589-14DL	MYE5N2D	DUP	11/13/24 00:28	4X for Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133416

Review By	Sarabjit Jaswal	Review On	11/14/2024 7:16:37 AM
Supervise By	Mohan Bera	Supervise On	11/14/2024 11:07:21 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

213	P4589-13LDL	MYE5N2L	SD	11/13/24 00:31	20X for Ba,Co,Ni,V	Jaswal	Confirms
214	P4589-15DL	MYE5N2S	MS	11/13/24 00:34	4X for Ba,Co,Ni,V	Jaswal	Confirms
215	P4589-13MSDL	MYE5N2SRE	MS	11/13/24 00:38	4X for Ba,Co,Ni,V	Jaswal	Confirms
216	P4588-14DL	MYE5K5	SAM	11/13/24 00:47	Not USe	Jaswal	Not Ok
217	P4588-15DL	MYE5K5D	DUP	11/13/24 00:50	Not USe	Jaswal	Not Ok
218	P4588-14LDL	MYE5K5L	SD	11/13/24 00:53	Not USe	Jaswal	Not Ok
219	P4588-16DL	MYE5K5S	MS	11/13/24 00:57	Not USe	Jaswal	Not Ok
220	P4588-14MSDL	MYE5K5SRE	MS	11/13/24 01:00	Not USe	Jaswal	Not Ok
221	CCV008	CCV008	CCV	11/13/24 01:03		Jaswal	OK
222	CCB008	CCB008	CCB	11/13/24 01:05		Jaswal	OK