## DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Labora	atory Name :	Alliance Technical Group LLC	al Group LLC Client :		Portal Partners	Tri-Ve	nture			
Project Location : NJ			Project N	umber :	9500000878					
Laboratory Sample ID(s): P4578 Sampling Date(s): 10/25/2024										
List DKQP Methods Used (e.g., 8260,8270, et Cetra) ,1030,1311,1311 ZHE, 6010D, 7196A, 7470A, 7471B, 8081B, 8082A, 8151A 8260-Low, 8260D, 8270E, 9012B, 9034, 9045D, NJEPH									A, 8151A,	
1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?					V	Yes		No	
1A	Were the method specified handling, preservation, and holding time requirements met?						Yes	$\overline{\checkmark}$	No	
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)					V	Yes		No	□ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?					V	Yes		No	
3	Were samples received at an appropriate temperature (4±2° C)?					$\overline{\mathbf{V}}$	Yes		No	□ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?						Yes	V	No	
5		ing limits specified or reference I to the laboratory prior to samp		ody or		V	Yes		No	
	b)Were these	reporting limits met?				$\overline{\mathbf{V}}$	Yes		No	□ N/A
6	results report	ytical method referenced in this ed for all constituents identified ne DKQP documents and/or sit	in the method-specifi			V	Yes		No	
7	Are project-sp	ecific matrix spikes and/or labo	atory duplicates inclu	ded in th	is data set?		Yes	$\overline{\mathbf{A}}$	No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."