



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4578				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4578	SAS No.:	P4578		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB55	Mercury	0.20	+/-0.20	U	0.20	CV	10/28/2024	11:14	LB133167	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4578			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4578	SAS No.:	P4578	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	10/28/2024	11:21	LB133167
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	10/28/2024	11:53	LB133167
CCB81	Mercury	0.20	+/-0.20	U	0.20	CV	10/28/2024	12:14	LB133167

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4578			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4578	SAS No.:	P4578	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB57	Mercury	0.20	+/-0.20	U	0.20	CV	10/29/2024	10:34	LB133185

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Contract: PORT06

Lab Code: CHEM

Case No.: P4578

SAS No.: P4578

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	10/29/2024	10:39	LB133185
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	10/29/2024	11:06	LB133185
CCB88	Mercury	0.20	+/-0.20	U	0.20	CV	10/29/2024	11:34	LB133185
CCB89	Mercury	0.20	+/-0.20	U	0.20	CV	10/29/2024	11:59	LB133185

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	10/28/2024	13:08	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	13:08	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	13:08	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	13:08	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	13:08	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	13:08	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	13:08	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	13:08	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	13:08	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	13:08	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	13:08	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	13:08	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	13:08	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	13:08	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	13:08	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	13:08	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	13:08	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	13:08	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	13:08	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	13:08	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	13:08	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	13:08	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	10/28/2024	13:39	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	13:39	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	13:39	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	13:39	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	13:39	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	13:39	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	13:39	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	13:39	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	13:39	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	13:39	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	13:39	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	13:39	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	13:39	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	13:39	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	13:39	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	13:39	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	13:39	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	13:39	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	13:39	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	13:39	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	13:39	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	13:39	LB133175
CCB02	Aluminum	100	+/-100	U	100	P	10/28/2024	14:30	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	14:30	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	14:30	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	14:30	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	14:30	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	14:30	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	14:30	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	14:30	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	14:30	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	14:30	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	14:30	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	14:30	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	14:30	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	14:30	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	14:30	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	14:30	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	14:30	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	14:30	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	14:30	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	14:30	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	14:30	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	14:30	LB133175
CCB03	Aluminum	100	+/-100	U	100	P	10/28/2024	15:20	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	15:20	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	15:20	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	15:20	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	15:20	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	15:20	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	15:20	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	15:20	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	15:20	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	15:20	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	15:20	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	15:20	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	15:20	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	15:20	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	15:20	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	15:20	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	15:20	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	15:20	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	15:20	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	15:20	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	15:20	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	15:20	LB133175
CCB04	Aluminum	100	+/-100	U	100	P	10/28/2024	16:10	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	16:10	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	16:10	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	16:10	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	16:10	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	16:10	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	16:10	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	16:10	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	16:10	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	16:10	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	16:10	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	16:10	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	16:10	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	16:10	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	16:10	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	16:10	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	16:10	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	16:10	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	16:10	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	16:10	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	16:10	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	16:10	LB133175
CCB05	Aluminum	100	+/-100	U	100	P	10/28/2024	17:01	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	17:01	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	17:01	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	17:01	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	17:01	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	17:01	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	17:01	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	17:01	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	17:01	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	17:01	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	17:01	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	17:01	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	17:01	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	17:01	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	17:01	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	17:01	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	17:01	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	17:01	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	17:01	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	17:01	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	17:01	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	17:01	LB133175
CCB06	Aluminum	100	+/-100	U	100	P	10/28/2024	17:54	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	17:54	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	17:54	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	17:54	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	17:54	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	17:54	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	17:54	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	17:54	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	17:54	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	17:54	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	17:54	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	17:54	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	17:54	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	17:54	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	17:54	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	17:54	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	17:54	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	17:54	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	17:54	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	17:54	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	17:54	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	17:54	LB133175
CCB07	Aluminum	100	+/-100	U	100	P	10/28/2024	18:47	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	18:47	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	18:47	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	18:47	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	18:47	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	18:47	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	18:47	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	18:47	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	18:47	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	18:47	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	18:47	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	18:47	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	18:47	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	18:47	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	18:47	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	18:47	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	18:47	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	18:47	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	18:47	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	18:47	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	18:47	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	18:47	LB133175
CCB08	Aluminum	100	+/-100	U	100	P	10/28/2024	19:37	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	19:37	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	19:37	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	19:37	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	19:37	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	19:37	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	19:37	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	19:37	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	19:37	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	19:37	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	19:37	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	19:37	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	19:37	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	19:37	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	19:37	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	19:37	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	19:37	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	19:37	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	19:37	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	19:37	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	19:37	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	19:37	LB133175
CCB09	Aluminum	100	+/-100	U	100	P	10/28/2024	20:26	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	20:26	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	20:26	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	20:26	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	20:26	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	20:26	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	20:26	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	20:26	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	20:26	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	20:26	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	20:26	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	20:26	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	20:26	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	20:26	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	20:26	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	20:26	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	20:26	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	20:26	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	20:26	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	20:26	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	20:26	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	20:26	LB133175
CCB10	Aluminum	100	+/-100	U	100	P	10/28/2024	21:25	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	21:25	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	21:25	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	21:25	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	21:25	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	21:25	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	21:25	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	21:25	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	21:25	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	21:25	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	21:25	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	21:25	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	21:25	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	21:25	LB133175
	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	21:25	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	21:25	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	21:25	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	21:25	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	21:25	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	21:25	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	21:25	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	21:25	LB133175
CCB11	Aluminum	100	+/-100	U	100	P	10/28/2024	21:43	LB133175
	Antimony	50.0	+/-50.0	U	50.0	P	10/28/2024	21:43	LB133175
	Arsenic	20.0	+/-20.0	U	20.0	P	10/28/2024	21:43	LB133175
	Barium	100	+/-100	U	100	P	10/28/2024	21:43	LB133175
	Beryllium	6.00	+/-6.00	U	6.00	P	10/28/2024	21:43	LB133175
	Cadmium	6.00	+/-6.00	U	6.00	P	10/28/2024	21:43	LB133175
	Calcium	2000	+/-2000	U	2000	P	10/28/2024	21:43	LB133175
	Chromium	10.0	+/-10.0	U	10.0	P	10/28/2024	21:43	LB133175
	Cobalt	30.0	+/-30.0	U	30.0	P	10/28/2024	21:43	LB133175
	Copper	20.0	+/-20.0	U	20.0	P	10/28/2024	21:43	LB133175
	Iron	100	+/-100	U	100	P	10/28/2024	21:43	LB133175
	Lead	12.0	+/-12.0	U	12.0	P	10/28/2024	21:43	LB133175
	Magnesium	2000	+/-2000	U	2000	P	10/28/2024	21:43	LB133175
	Manganese	20.0	+/-20.0	U	20.0	P	10/28/2024	21:43	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract:	PORT06	Lab Code:	CHEM	Case No.:	P4578	SAS No.:	P4578		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	40.0	+/-40.0	U	40.0	P	10/28/2024	21:43	LB133175
	Potassium	2000	+/-2000	U	2000	P	10/28/2024	21:43	LB133175
	Selenium	20.0	+/-20.0	U	20.0	P	10/28/2024	21:43	LB133175
	Silver	10.0	+/-10.0	U	10.0	P	10/28/2024	21:43	LB133175
	Sodium	2000	+/-2000	U	2000	P	10/28/2024	21:43	LB133175
	Thallium	40.0	+/-40.0	U	40.0	P	10/28/2024	21:43	LB133175
	Vanadium	40.0	+/-40.0	U	40.0	P	10/28/2024	21:43	LB133175
	Zinc	40.0	+/-40.0	U	40.0	P	10/28/2024	21:43	LB133175

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	10/29/2024	12:35	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	12:35	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	12:35	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	12:35	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	12:35	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	12:35	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	12:35	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	12:35	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	12:35	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	12:35	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	12:35	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	12:35	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	12:35	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	12:35	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	12:35	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	12:35	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	12:35	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	12:35	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	12:35	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	12:35	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	12:35	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	12:35	LB133198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	10/29/2024	13:06	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	13:06	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	13:06	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	13:06	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	13:06	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	13:06	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	13:06	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	13:06	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	13:06	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	13:06	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	13:06	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	13:06	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	13:06	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	13:06	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	13:06	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	13:06	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	13:06	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	13:06	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	13:06	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	13:06	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	13:06	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	13:06	LB133198
CCB02	Aluminum	100	+/-100	U	100	P	10/29/2024	14:15	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	14:15	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	14:15	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	14:15	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	14:15	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	14:15	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	14:15	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	14:15	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	14:15	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	14:15	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	14:15	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	14:15	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	14:15	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	14:15	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	14:15	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	14:15	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	14:15	LB133198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Contract: PORT06

Lab Code: CHEM

Case No.: P4578

SAS No.: P4578

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	14:15	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	14:15	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	14:15	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	14:15	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	14:15	LB133198
CCB03	Aluminum	100	+/-100	U	100	P	10/29/2024	15:06	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	15:06	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	15:06	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	15:06	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	15:06	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	15:06	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	15:06	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	15:06	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	15:06	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	15:06	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	15:06	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	15:06	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	15:06	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	15:06	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	15:06	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	15:06	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	15:06	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	15:06	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	15:06	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	15:06	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	15:06	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	15:06	LB133198
CCB04	Aluminum	100	+/-100	U	100	P	10/29/2024	16:03	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	16:03	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	16:03	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	16:03	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	16:03	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	16:03	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	16:03	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	16:03	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	16:03	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	16:03	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	16:03	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	16:03	LB133198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	16:03	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	16:03	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	16:03	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	16:03	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	16:03	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	16:03	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	16:03	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	16:03	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	16:03	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	16:03	LB133198
CCB05	Aluminum	100	+/-100	U	100	P	10/29/2024	16:57	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	16:57	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	16:57	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	16:57	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	16:57	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	16:57	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	16:57	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	16:57	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	16:57	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	16:57	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	16:57	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	16:57	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	16:57	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	16:57	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	16:57	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	16:57	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	16:57	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	16:57	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	16:57	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	16:57	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	16:57	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	16:57	LB133198
CCB06	Aluminum	100	+/-100	U	100	P	10/29/2024	17:58	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	17:58	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	17:58	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	17:58	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	17:58	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	17:58	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	17:58	LB133198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	17:58	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	17:58	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	17:58	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	17:58	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	17:58	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	17:58	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	17:58	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	17:58	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	17:58	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	17:58	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	17:58	LB133198
	Sodium	561	+/-2000	J	2000	P	10/29/2024	17:58	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	17:58	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	17:58	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	17:58	LB133198
CCB07	Aluminum	100	+/-100	U	100	P	10/29/2024	18:24	LB133198
	Antimony	50.0	+/-50.0	U	50.0	P	10/29/2024	18:24	LB133198
	Arsenic	20.0	+/-20.0	U	20.0	P	10/29/2024	18:24	LB133198
	Barium	100	+/-100	U	100	P	10/29/2024	18:24	LB133198
	Beryllium	6.00	+/-6.00	U	6.00	P	10/29/2024	18:24	LB133198
	Cadmium	6.00	+/-6.00	U	6.00	P	10/29/2024	18:24	LB133198
	Calcium	2000	+/-2000	U	2000	P	10/29/2024	18:24	LB133198
	Chromium	10.0	+/-10.0	U	10.0	P	10/29/2024	18:24	LB133198
	Cobalt	30.0	+/-30.0	U	30.0	P	10/29/2024	18:24	LB133198
	Copper	20.0	+/-20.0	U	20.0	P	10/29/2024	18:24	LB133198
	Iron	100	+/-100	U	100	P	10/29/2024	18:24	LB133198
	Lead	12.0	+/-12.0	U	12.0	P	10/29/2024	18:24	LB133198
	Magnesium	2000	+/-2000	U	2000	P	10/29/2024	18:24	LB133198
	Manganese	20.0	+/-20.0	U	20.0	P	10/29/2024	18:24	LB133198
	Nickel	40.0	+/-40.0	U	40.0	P	10/29/2024	18:24	LB133198
	Potassium	2000	+/-2000	U	2000	P	10/29/2024	18:24	LB133198
	Selenium	20.0	+/-20.0	U	20.0	P	10/29/2024	18:24	LB133198
	Silver	10.0	+/-10.0	U	10.0	P	10/29/2024	18:24	LB133198
	Sodium	2000	+/-2000	U	2000	P	10/29/2024	18:24	LB133198
	Thallium	40.0	+/-40.0	U	40.0	P	10/29/2024	18:24	LB133198
	Vanadium	40.0	+/-40.0	U	40.0	P	10/29/2024	18:24	LB133198
	Zinc	40.0	+/-40.0	U	40.0	P	10/29/2024	18:24	LB133198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:16	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:16	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:50	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:50	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
CCB02	Aluminum	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	15:33	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	15:33	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
CCB03	Aluminum	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	16:28	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	16:28	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
CCB04	Aluminum	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	17:22	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	17:22	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
CCB05	Aluminum	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:08	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:08	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
CCB06	Aluminum	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:59	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:59	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
CCB07	Aluminum	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	19:51	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	19:51	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Sodium	941	+/-2000	J	2000	P	11/08/2024	19:51	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
CCB08	Aluminum	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	20:57	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	20:57	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Sodium	720	+/-2000	J	2000	P	11/08/2024	20:57	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
CCB09	Aluminum	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	21:55	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	21:55	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4578							
Contract: PORT06	Lab Code: CHEM	Case No.: P4578	SAS No.: P4578						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
CCB10	Aluminum	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	23:35	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	23:35	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Sodium	742	+/-2000	J	2000	P	11/08/2024	23:35	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365