

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164472BL		WATER		Batch Number:	PB164472		Prep Date:	10/25/2024	
	Mercury	0.20	<0.20	U	0.20	CV	10/28/2024	11:32	LB133167
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164487BL		SOLID		Batch Number:	PB164487		Prep Date:	10/28/2024	
	Mercury	0.013	<0.013	U	0.013	CV	10/29/2024	10:48	LB133185

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164415BL	SOLID			Batch Number:	PB164415		Prep Date:	10/25/2024	
	Aluminum	4.22	<4.22	U	4.22	P	10/28/2024	21:08	LB133175
	Antimony	2.11	<2.11	U	2.11	P	10/28/2024	21:08	LB133175
	Arsenic	0.84	<0.84	U	0.84	P	10/28/2024	21:08	LB133175
	Barium	4.22	<4.22	U	4.22	P	10/28/2024	21:08	LB133175
	Beryllium	0.25	<0.25	U	0.25	P	10/28/2024	21:08	LB133175
	Cadmium	0.25	<0.25	U	0.25	P	10/28/2024	21:08	LB133175
	Calcium	84.4	<84.4	U	84.4	P	10/28/2024	21:08	LB133175
	Chromium	0.42	<0.42	U	0.42	P	10/28/2024	21:08	LB133175
	Cobalt	1.27	<1.27	U	1.27	P	10/28/2024	21:08	LB133175
	Copper	0.84	<0.84	U	0.84	P	10/28/2024	21:08	LB133175
	Iron	4.22	<4.22	U	4.22	P	10/28/2024	21:08	LB133175
	Lead	0.51	<0.51	U	0.51	P	10/28/2024	21:08	LB133175
	Magnesium	84.4	<84.4	U	84.4	P	10/28/2024	21:08	LB133175
	Manganese	0.84	<0.84	U	0.84	P	10/28/2024	21:08	LB133175
	Nickel	1.69	<1.69	U	1.69	P	10/28/2024	21:08	LB133175
	Potassium	84.4	<84.4	U	84.4	P	10/28/2024	21:08	LB133175
	Selenium	0.84	<0.84	U	0.84	P	10/28/2024	21:08	LB133175
	Silver	0.42	<0.42	U	0.42	P	10/28/2024	21:08	LB133175
	Sodium	84.4	<84.4	U	84.4	P	10/28/2024	21:08	LB133175
	Thallium	1.69	<1.69	U	1.69	P	10/28/2024	21:08	LB133175
	Vanadium	1.69	<1.69	U	1.69	P	10/28/2024	21:08	LB133175
	Zinc	1.69	<1.69	U	1.69	P	10/28/2024	21:08	LB133175
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164438BL	WATER			Batch Number:	PB164438		Prep Date:	10/26/2024	
	Aluminum	50.0	<50.0	U	50.0	P	10/29/2024	18:07	LB133198
	Antimony	25.0	<25.0	U	25.0	P	10/29/2024	18:07	LB133198
	Arsenic	10.0	<10.0	U	10.0	P	10/29/2024	18:07	LB133198
	Barium	50.0	<50.0	U	50.0	P	10/29/2024	18:07	LB133198
	Beryllium	3.00	<3.00	U	3.00	P	10/29/2024	18:07	LB133198
	Cadmium	3.00	<3.00	U	3.00	P	10/29/2024	18:07	LB133198
	Calcium	1000	<1000	U	1000	P	10/29/2024	18:07	LB133198
	Chromium	5.00	<5.00	U	5.00	P	10/29/2024	18:07	LB133198
	Cobalt	15.0	<15.0	U	15.0	P	10/29/2024	18:07	LB133198
	Copper	10.0	<10.0	U	10.0	P	10/29/2024	18:07	LB133198
	Iron	50.0	<50.0	U	50.0	P	10/29/2024	18:07	LB133198
	Lead	6.00	<6.00	U	6.00	P	10/29/2024	18:07	LB133198
	Magnesium	1000	<1000	U	1000	P	10/29/2024	18:07	LB133198
	Manganese	10.0	<10.0	U	10.0	P	10/29/2024	18:07	LB133198
	Nickel	20.0	<20.0	U	20.0	P	10/29/2024	18:07	LB133198

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Instrument: P4

Potassium	1000	<1000	U	1000	P	10/29/2024	18:07	LB133198
Selenium	10.0	<10.0	U	10.0	P	10/29/2024	18:07	LB133198
Silver	5.00	<5.00	U	5.00	P	10/29/2024	18:07	LB133198
Sodium	386	<1000	J	1000	P	10/29/2024	18:07	LB133198
Thallium	20.0	<20.0	U	20.0	P	10/29/2024	18:07	LB133198
Vanadium	20.0	<20.0	U	20.0	P	10/29/2024	18:07	LB133198
Zinc	20.0	<20.0	U	20.0	P	10/29/2024	18:07	LB133198