


LB133181

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	-86	0.000	-86	0				2
0.2	0.200	0.203	0.003	582	0.0 %	582					-1
2.5	2.500	2.467	-0.033	8537	0.0 %	8537					-1
5.0	5.000	4.957	-0.043	17287	0.0 %	17287					2
7.5	7.500	7.632	0.132	26686	0.0 %	26686					-1
10.0	10.000	9.928	-0.072	34752	0.0 %	34752					-1

LB133181 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Oct 2024 07:38:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	-86	-	0.0000	7470A	PPB	29 Oct 2024 08:01:13	S	Std
0.2 - 1	50.2	582	-	0.2000	7470A	PPB	29 Oct 2024 08:03:30	S	Std
2.5 - 1	52.5	8537	-	2.5000	7470A	PPB	29 Oct 2024 08:05:47	S	Std
5.0 - 1	55	17287	-	5.0000	7470A	PPB	29 Oct 2024 08:08:03	S	Std
7.5 - 1	57.5	26686	-	7.5000	7470A	PPB	29 Oct 2024 08:10:20	S	Std
10.0 - 1	510	34752	-	10.0000	7470A	PPB	29 Oct 2024 08:12:38	S	Std
ICV56 - 1	ICV56	13482	3.8743	-	7470A	PPB	29 Oct 2024 08:15:40	U	SMPL
ICB56 - 1	ICB56	-380	-0.0709	-	7470A	PPB	29 Oct 2024 08:17:54	U	SMPL
CCV82 - 1	CCV82	17569	5.0375	-	7470A	PPB	29 Oct 2024 08:20:10	U	SMPL
CCB82 - 1	CCB82	-401	-0.0768	-	7470A	PPB	29 Oct 2024 08:22:25	U	SMPL
CRA - 1	CRA	502	0.1802	-	7470A	PPB	29 Oct 2024 08:24:42	U	SMPL
HighStd - 1	HighStd	35409	10.1148	-	7470A	PPB	29 Oct 2024 08:26:57	U	SMPL
ChkStd - 1	ChkStd	24676	7.0602	-	7470A	PPB	29 Oct 2024 08:29:12	U	SMPL
PB164477BL - 1	PBW	-510	-0.1079	-	7470A	PPB	29 Oct 2024 08:31:30	U	SMPL
PB164477BS - 1	LCSW	13787	3.9611	-	7470A	PPB	29 Oct 2024 08:33:49	U	SMPL
P4545-02 - 1	VNJ-215	-357	-0.0643	-	7470A	PPB	29 Oct 2024 08:36:04	U	SMPL
P4547-04 - 1	BP-F-21	-124	0.0020	-	7470A	PPB	29 Oct 2024 08:38:22	U	SMPL
P4547-08 - 1	BP-F-20	-42	0.0253	-	7470A	PPB	29 Oct 2024 08:40:39	U	SMPL
P4561-04 - 1	BP-F-19	-54	0.0219	-	7470A	PPB	29 Oct 2024 08:42:56	U	SMPL
P4561-08 - 1	BP-F-18	-61	0.0199	-	7470A	PPB	29 Oct 2024 08:45:11	U	SMPL
CCV83 - 1	CCV83	17105	4.9054	-	7470A	PPB	29 Oct 2024 08:47:26	U	SMPL
CCB83 - 1	CCB83	-353	-0.0632	-	7470A	PPB	29 Oct 2024 08:49:41	U	SMPL
P4567-04 - 1	WC-1	-133	-0.0006	-	7470A	PPB	29 Oct 2024 08:51:58	U	SMPL
P4567-08 - 1	WC-2	-102	0.0083	-	7470A	PPB	29 Oct 2024 08:54:13	U	SMPL
P4567-12 - 1	WC-3	-106	0.0071	-	7470A	PPB	29 Oct 2024 08:56:28	U	SMPL
P4574-03 - 1	GRAVEL-1	-99	0.0091	-	7470A	PPB	29 Oct 2024 08:58:44	U	SMPL
P4574-06 - 1	GRAVEL-2	-95	0.0103	-	7470A	PPB	29 Oct 2024 09:00:59	U	SMPL
P4574-08 - 1	CONCRETE-1	-79	0.0148	-	7470A	PPB	29 Oct 2024 09:03:15	U	SMPL
P4574-10 - 1	CONCRETE-2	-63	0.0194	-	7470A	PPB	29 Oct 2024 09:05:31	U	SMPL
P4574-12 - 1	CONCRETE-3	-73	0.0165	-	7470A	PPB	29 Oct 2024 09:07:48	U	SMPL
P4578-06 - 1	WB-305-BOT	-64	0.0191	-	7470A	PPB	29 Oct 2024 09:10:04	U	SMPL
P4578-08 - 1	WB-305-BOT-1	-80	0.0145	-	7470A	PPB	29 Oct 2024 09:12:20	U	SMPL
CCV84 - 1	CCV84	15986	4.5870	-	7470A	PPB	29 Oct 2024 09:14:35	U	SMPL
CCB84 - 1	CCB84	-339	-0.0592	-	7470A	PPB	29 Oct 2024 09:16:51	U	SMPL
P4578-08DUP - 1	WB-305-BOT-1DUP	-57	0.0211	-	7470A	PPB	29 Oct 2024 09:19:08	U	SMPL
P4578-08MS - 1	WB-305-BOT-1MS	13480	3.8737	-	7470A	PPB	29 Oct 2024 09:26:16	U	SMPL
P4578-08MSD - 1	WB-305-BOT-1MSD	12954	3.7240	-	7470A	PPB	29 Oct 2024 09:28:30	U	SMPL
PB164393TB - 1	PB164393TB	-356	-0.0640	-	7470A	PPB	29 Oct 2024 09:30:48	U	SMPL
P4578-08LX5 - 1		-191	-0.0171	-	7470A	PPB	29 Oct 2024 09:33:05	U	SMPL
P4578-08A - 1		13720	3.9421	-	7470A	PPB	29 Oct 2024 09:35:21	U	SMPL
CCV85 - 1	CCV85	17743	5.0870	-	7470A	PPB	29 Oct 2024 09:37:37	U	SMPL
CCB85 - 1	CCB85	-362	-0.0657	-	7470A	PPB	29 Oct 2024 09:39:55	U	SMPL