



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** P4578
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4578 **SAS No.:** P4578
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.1	20.0	96	40 - 160	P	10/28/2024	13:12	LB133175
	Barium	93.7	100	94	40 - 160	P	10/28/2024	13:12	LB133175
	Cadmium	5.55	6.0	92	40 - 160	P	10/28/2024	13:12	LB133175
	Chromium	9.13	10.0	91	40 - 160	P	10/28/2024	13:12	LB133175
	Lead	11.5	12.0	96	40 - 160	P	10/28/2024	13:12	LB133175
	Selenium	17.6	20.0	88	40 - 160	P	10/28/2024	13:12	LB133175
	Silver	11.0	10.0	110	40 - 160	P	10/28/2024	13:12	LB133175
CRA	Mercury	0.18	0.2	90	40 - 160	CV	10/29/2024	08:24	LB133181