

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P4578
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4578 **SAS No.:** P4578
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.59			-20	20	10/28/2024	13:16	LB133175
	Barium	2.22	6.0	37	-94	106	10/28/2024	13:16	LB133175
	Cadmium	6.97	1.0	697	-5	7	10/28/2024	13:16	LB133175
	Chromium	56.5	52.0	109	42	62	10/28/2024	13:16	LB133175
	Lead	8.36			-12	12	10/28/2024	13:16	LB133175
	Selenium	-9.09			-20	20	10/28/2024	13:16	LB133175
	Silver	3.99			-10	10	10/28/2024	13:16	LB133175
ICSAB01	Arsenic	107	104	103	88.4	120	10/28/2024	13:21	LB133175
	Barium	472	537	88	437	637	10/28/2024	13:21	LB133175
	Cadmium	967	972	100	826	1120	10/28/2024	13:21	LB133175
	Chromium	556	542	103	460	624	10/28/2024	13:21	LB133175
	Lead	55.7	49.0	114	37	61	10/28/2024	13:21	LB133175
	Selenium	39.8	46.0	86	26	66	10/28/2024	13:21	LB133175
	Silver	212	201	106	170	232	10/28/2024	13:21	LB133175