

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4578
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4578 **SAS No.:** P4578
Matrix: Water **Sample ID:** P4578-08 **Client ID:** WB-305-BOT-1DUP
Percent Solids for Sample: NA **Duplicate ID** P4578-08DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	492	J	472	J	4		P
Cadmium	ug/L	20	1.08	J	1.39	J	25		P
Chromium	ug/L	20	42.4	J	41.7	J	2		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4578
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4578 **SAS No.:** P4578
Matrix: Water **Sample ID:** P4578-08MS **Client ID:** WB-305-BOT-1MSD
Percent Solids for Sample: NA **Duplicate ID** P4578-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3670		3710		1		P
Barium	ug/L	20	1330		1330		0		P
Cadmium	ug/L	20	874		874		0		P
Chromium	ug/L	20	1910		1920		1		P
Lead	ug/L	20	4140		4140		0		P
Mercury	ug/L	20	38.7		37.2		4		CV
Selenium	ug/L	20	9380		9510		1		P
Silver	ug/L	20	373		381		2		P