

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4578
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4578 **SAS No.:** P4578

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164453BS							
Arsenic	ug/L	4000	3870		97	80 - 120	P
Barium	ug/L	1000	945		94	80 - 120	P
Cadmium	ug/L	1000	943		94	80 - 120	P
Chromium	ug/L	2000	2000		100	80 - 120	P
Lead	ug/L	5000	4800		96	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	399		105	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4578

Contract: PORT06

Lab Code: CHEM

Case No.: P4578

SAS No.: P4578

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164477BS Mercury	ug/L	4.0	3.96		99	80 - 120	CV