

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4578 **SAS No :** P4578 **SDG No:** P4578
Sample No : PB164525BS **Datafile:** FC067569.D
Client ID : PB164525BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10.0	5.21	52		(40-140)
Aliphatic C12-C16	6.7	5.06	76		(40-140)
Aliphatic C16-C21	10.0	8.68	87		(40-140)
Aliphatic C21-C28	13.3	11.8	88		(40-140)
Aliphatic C28-C40	20.0	18.9	94		(40-140)