

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4578 **SAS No :** P4578 **SDG No:** P4578
Sample No : PB164532BS **Datafile:** FE050918.D
Client ID : PB164532BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	167	56		(40-140)
Aliphatic C12-C16	200	155	78		(40-140)
Aliphatic C16-C21	300	267	89		(40-140)
Aliphatic C21-C28	400	363	91		(40-140)
Aliphatic C28-C40	600	607	101		(40-140)