

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4578 **SAS No :** P4578 **SDG No:** P4578
Sample No : PB164532BSD **Datafile:** FE050919.D
Client ID : PB164532BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	165	55		1.3 (40-140)	50
Aliphatic C12-C16	200	153	76		1.3 (40-140)	50
Aliphatic C16-C21	300	265	88		0.752 (40-140)	50
Aliphatic C21-C28	400	360	90		0.885 (40-140)	50
Aliphatic C28-C40	600	612	102		0.98 (40-140)	50