

**WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** Portal Partners Tri-Venture  
**Lab Code:** CHEM **Cas No:** P4578 **SAS No :** P4578 **SDG No:** P4578  
**Sample No :** PB164532BS **Datafile:** FG014589.D  
**Client ID :** PB164532BS

| COMPOUND         | SPIKE<br>ADDED<br>ug/l | LCS/LCSD<br>CONCENTRATION<br>ug/l | % REC | Qual | QC LIMITS |
|------------------|------------------------|-----------------------------------|-------|------|-----------|
| Aromatic C10-C12 | 200                    | 182                               | 91    |      | (40-140)  |
| Aromatic C12-C16 | 300                    | 300                               | 100   |      | (40-140)  |
| Aromatic C16-C21 | 500                    | 589                               | 117   |      | (40-140)  |
| Aromatic C21-C36 | 800                    | 889                               | 111   |      | (40-140)  |