

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Continuing Calib Date: 10/28/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 21:45 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1242-2 (2)	5.54	5.54	5.44	5.64	0.00
Aroclor-1242-3 (3)	5.60	5.61	5.51	5.71	0.01
Aroclor-1242-4 (4)	5.70	5.70	5.60	5.80	0.00
Aroclor-1242-5 (5)	6.44	6.44	6.34	6.54	0.01
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.06	10.06	9.96	10.16	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Continuing Calib Date: 10/28/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 21:45 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.75	4.73	4.63	4.83	-0.02
Aroclor-1242-2 (2)	4.77	4.75	4.65	4.85	-0.02
Aroclor-1242-3 (3)	4.95	4.92	4.82	5.02	-0.03
Aroclor-1242-4 (4)	4.99	5.00	4.90	5.10	0.01
Aroclor-1242-5 (5)	5.55	5.53	5.43	5.63	-0.02
Tetrachloro-m-xylene	3.67	3.64	3.54	3.74	-0.03
Decachlorobiphenyl	8.66	8.64	8.54	8.74	-0.02