

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4578 **SAS No.:** P4578 **SDG NO.:** P4578

Continuing Calib Date: 10/29/2024 **Initial Calibration Date(s):** 10/08/2024 10/08/2024

Continuing Calib Time: 10:11 **Initial Calibration Time(s):** 16:30 23:46

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.92	5.92	5.82	6.02	0.01
Aroclor-1242-2 (2)	5.94	5.94	5.84	6.04	0.00
Aroclor-1242-3 (3)	6.00	6.00	5.90	6.10	0.00
Aroclor-1242-4 (4)	6.10	6.10	6.00	6.20	0.00
Aroclor-1242-5 (5)	6.83	6.84	6.74	6.94	0.01
Tetrachloro-m-xylene	4.75	4.75	4.65	4.85	0.00
Decachlorobiphenyl	10.66	10.67	10.57	10.77	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Continuing Calib Date: 10/29/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 10:11 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.15	5.16	5.06	5.26	0.01
Aroclor-1242-2 (2)	5.17	5.18	5.08	5.28	0.01
Aroclor-1242-3 (3)	5.35	5.36	5.26	5.46	0.01
Aroclor-1242-4 (4)	5.44	5.45	5.35	5.55	0.01
Aroclor-1242-5 (5)	5.97	5.98	5.88	6.08	0.01
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.20	9.22	9.12	9.32	0.02