

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107435.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.519	5.423	5.623	480.930	500.000	-3.8
Aroclor-1016-2	5.542	5.445	5.645	469.040	500.000	-6.2
Aroclor-1016-3	5.604	5.507	5.707	469.220	500.000	-6.2
Aroclor-1016-4	5.701	5.604	5.804	488.600	500.000	-2.3
Aroclor-1016-5	5.996	5.899	6.099	484.120	500.000	-3.2
Aroclor-1260-1	7.123	7.026	7.226	459.280	500.000	-8.1
Aroclor-1260-2	7.379	7.282	7.482	476.130	500.000	-4.8
Aroclor-1260-3	7.741	7.643	7.843	470.430	500.000	-5.9
Aroclor-1260-4	7.965	7.868	8.068	465.450	500.000	-6.9
Aroclor-1260-5	8.277	8.181	8.381	480.940	500.000	-3.8
Decachlorobiphenyl	10.055	9.958	10.158	47.620	50.000	-4.8
Tetrachloro-m-xylene	4.371	4.274	4.474	47.610	50.000	-4.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107435.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.626	4.826	523.860	500.000	4.8
Aroclor-1016-2	4.743	4.645	4.845	530.430	500.000	6.1
Aroclor-1016-3	4.919	4.821	5.021	508.190	500.000	1.6
Aroclor-1016-4	4.960	4.862	5.062	491.260	500.000	-1.7
Aroclor-1016-5	5.173	5.076	5.276	531.290	500.000	6.3
Aroclor-1260-1	6.204	6.107	6.307	493.890	500.000	-1.2
Aroclor-1260-2	6.390	6.294	6.494	525.730	500.000	5.1
Aroclor-1260-3	6.544	6.448	6.648	507.540	500.000	1.5
Aroclor-1260-4	7.014	6.918	7.118	492.640	500.000	-1.5
Aroclor-1260-5	7.255	7.159	7.359	527.570	500.000	5.5
Decachlorobiphenyl	8.634	8.539	8.739	48.670	50.000	-2.7
Tetrachloro-m-xylene	3.643	3.545	3.745	51.710	50.000	3.4