

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107451.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.519	5.423	5.623	419.080	500.000	-16.2
Aroclor-1016-2	5.542	5.445	5.645	408.530	500.000	-18.3
Aroclor-1016-3	5.605	5.507	5.707	396.930	500.000	-20.6
Aroclor-1016-4	5.701	5.604	5.804	414.160	500.000	-17.2
Aroclor-1016-5	5.995	5.899	6.099	444.570	500.000	-11.1
Aroclor-1260-1	7.122	7.026	7.226	416.700	500.000	-16.7
Aroclor-1260-2	7.379	7.282	7.482	413.910	500.000	-17.2
Aroclor-1260-3	7.741	7.643	7.843	409.630	500.000	-18.1
Aroclor-1260-4	7.964	7.868	8.068	421.520	500.000	-15.7
Aroclor-1260-5	8.278	8.181	8.381	434.130	500.000	-13.2
Decachlorobiphenyl	10.055	9.958	10.158	44.130	50.000	-11.7
Tetrachloro-m-xylene	4.371	4.274	4.474	49.810	50.000	-0.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/28/2024

Lab Sample No.: AR1660CCC500 Data File : PO107451.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	532.630	500.000	6.5
Aroclor-1016-2	4.744	4.645	4.845	543.130	500.000	8.6
Aroclor-1016-3	4.919	4.821	5.021	518.800	500.000	3.8
Aroclor-1016-4	4.961	4.862	5.062	498.430	500.000	-0.3
Aroclor-1016-5	5.174	5.076	5.276	521.590	500.000	4.3
Aroclor-1260-1	6.204	6.107	6.307	468.750	500.000	-6.3
Aroclor-1260-2	6.391	6.294	6.494	491.530	500.000	-1.7
Aroclor-1260-3	6.544	6.448	6.648	461.820	500.000	-7.6
Aroclor-1260-4	7.014	6.918	7.118	430.720	500.000	-13.9
Aroclor-1260-5	7.256	7.159	7.359	448.970	500.000	-10.2
Decachlorobiphenyl	8.634	8.539	8.739	39.290	50.000	-21.4
Tetrachloro-m-xylene	3.644	3.545	3.745	52.350	50.000	4.7