

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/29/2024

Lab Sample No.: AR1660CCC500 Data File : PP068299.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.915	5.819	6.019	560.040	500.000	12.0
Aroclor-1016-2	5.938	5.842	6.042	551.340	500.000	10.3
Aroclor-1016-3	6.000	5.905	6.105	550.550	500.000	10.1
Aroclor-1016-4	6.099	6.003	6.203	544.740	500.000	8.9
Aroclor-1016-5	6.393	6.298	6.498	532.370	500.000	6.5
Aroclor-1260-1	7.516	7.422	7.622	510.920	500.000	2.2
Aroclor-1260-2	7.769	7.675	7.875	502.770	500.000	0.6
Aroclor-1260-3	8.130	8.037	8.237	518.960	500.000	3.8
Aroclor-1260-4	8.367	8.275	8.475	514.860	500.000	3.0
Aroclor-1260-5	8.704	8.612	8.812	516.390	500.000	3.3
Decachlorobiphenyl	10.657	10.569	10.769	56.100	50.000	12.2
Tetrachloro-m-xylene	4.751	4.655	4.855	56.350	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/29/2024

Lab Sample No.: AR1660CCC500 Data File : PP068299.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.153	5.061	5.261	503.500	500.000	0.7
Aroclor-1016-2	5.173	5.081	5.281	499.880	500.000	0.0
Aroclor-1016-3	5.354	5.262	5.462	524.560	500.000	4.9
Aroclor-1016-4	5.393	5.301	5.501	523.430	500.000	4.7
Aroclor-1016-5	5.612	5.520	5.720	508.770	500.000	1.8
Aroclor-1260-1	6.654	6.564	6.764	493.590	500.000	-1.3
Aroclor-1260-2	6.840	6.750	6.950	495.080	500.000	-1.0
Aroclor-1260-3	6.998	6.908	7.108	493.360	500.000	-1.3
Aroclor-1260-4	7.472	7.383	7.583	496.630	500.000	-0.7
Aroclor-1260-5	7.711	7.622	7.822	495.530	500.000	-0.9
Decachlorobiphenyl	9.202	9.119	9.319	52.510	50.000	5.0
Tetrachloro-m-xylene	4.047	3.952	4.152	50.920	50.000	1.8