

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/29/2024

Lab Sample No.: AR1660CCC500 Data File : PP068312.D Time Analyzed: 17:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.925	5.819	6.019	545.450	500.000	9.1
Aroclor-1016-2	5.946	5.842	6.042	552.100	500.000	10.4
Aroclor-1016-3	6.010	5.905	6.105	536.070	500.000	7.2
Aroclor-1016-4	6.108	6.003	6.203	518.640	500.000	3.7
Aroclor-1016-5	6.402	6.298	6.498	547.610	500.000	9.5
Aroclor-1260-1	7.524	7.422	7.622	509.830	500.000	2.0
Aroclor-1260-2	7.778	7.675	7.875	513.410	500.000	2.7
Aroclor-1260-3	8.139	8.037	8.237	535.650	500.000	7.1
Aroclor-1260-4	8.377	8.275	8.475	524.710	500.000	4.9
Aroclor-1260-5	8.713	8.612	8.812	533.880	500.000	6.8
Decachlorobiphenyl	10.669	10.569	10.769	58.530	50.000	17.1
Tetrachloro-m-xylene	4.761	4.655	4.855	58.210	50.000	16.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/29/2024

Lab Sample No.: AR1660CCC500 Data File : PP068312.D Time Analyzed: 17:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	517.750	500.000	3.6
Aroclor-1016-2	5.179	5.081	5.281	530.920	500.000	6.2
Aroclor-1016-3	5.359	5.262	5.462	531.030	500.000	6.2
Aroclor-1016-4	5.399	5.301	5.501	530.080	500.000	6.0
Aroclor-1016-5	5.617	5.520	5.720	548.700	500.000	9.7
Aroclor-1260-1	6.660	6.564	6.764	507.760	500.000	1.6
Aroclor-1260-2	6.847	6.750	6.950	513.820	500.000	2.8
Aroclor-1260-3	7.004	6.908	7.108	513.220	500.000	2.6
Aroclor-1260-4	7.478	7.383	7.583	516.080	500.000	3.2
Aroclor-1260-5	7.716	7.622	7.822	529.410	500.000	5.9
Decachlorobiphenyl	9.208	9.119	9.319	58.550	50.000	17.1
Tetrachloro-m-xylene	4.053	3.952	4.152	52.210	50.000	4.4