



Contract:	<u>PORT06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4578</u>	SAS No.:	<u>P4578</u>	SDG NO.:	<u>P4578</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>11/06/2024</u>	<u>11/06/2024</u>		
		Calibration Times:		<u>09:48</u>	<u>11:24</u>		

LAB FILE ID:	RT 200 =	<u>PS028253.D</u>	RT 500 =	<u>PS028254.D</u>
RT 750 =	<u>PS028255.D</u>	RT 1000 =	<u>PS028256.D</u>	RT 1500 =
			<u>PS028257.D</u>	

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Instrument ID: ECD_S Calibration Date(s): 11/06/2024 11/06/2024

Calibration Times: 09:48 11:24

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028253.D</u>	RT 500 = <u>PS028254.D</u>
	RT 750 = <u>PS028255.D</u>	RT 1000 = <u>PS028256.D</u>
		RT 1500 = <u>PS028257.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.92	9.92	9.92	9.92	9.92	9.92	9.82	10.02
2,4-D	9.02	9.02	9.02	9.02	9.02	9.02	8.92	9.12
2,4-DCAA	7.77	7.77	7.77	7.77	7.77	7.76	7.66	7.86