

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** PORT06

**Lab Code:** CHEM      **Case No.:** P4578      **SAS No.:** P4578      **SDG NO.:** P4578

**Continuing Calib Date:** 11/07/2024      **Initial Calibration Date(s):** 11/06/2024      11/06/2024

**Continuing Calib Time:** 20:41      **Initial Calibration Time(s):** 09:48      11:24

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.25       | 7.26      | 7.16      | 7.36 | 0.01       |
| 2,4-D            | 8.38       | 8.39      | 8.29      | 8.49 | 0.01       |
| 2,4,5-TP(Silvex) | 9.27       | 9.28      | 9.18      | 9.38 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** PORT06

**Lab Code:** CHEM      **Case No.:** P4578      **SAS No.:** P4578      **SDG NO.:** P4578

**Continuing Calib Date:** 11/07/2024      **Initial Calibration Date(s):** 11/06/2024      11/06/2024

**Continuing Calib Time:** 20:41      **Initial Calibration Time(s):** 09:48      11:24

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| 2,4-DCAA         | 7.76       | 7.77      | 7.67      | 7.87  | 0.01       |
| 2,4-D            | 9.01       | 9.02      | 8.92      | 9.12  | 0.01       |
| 2,4,5-TP(Silvex) | 9.91       | 9.92      | 9.82      | 10.02 | 0.01       |