

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL01 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028260.D Time Analyzed: 12:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.277	9.177	9.377	725.700	712.500	1.9
2,4-D	8.394	8.294	8.494	710.400	705.000	0.8
2,4-DCAA	7.262	7.162	7.362	757.550	750.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL01 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028260.D Time Analyzed: 12:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.922	9.823	10.023	729.510	712.500	2.4
2,4-D	9.015	8.915	9.115	713.640	705.000	1.2
2,4-DCAA	7.765	7.665	7.865	759.640	750.000	1.3