

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028272.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.276	9.177	9.377	748.270	712.500	5.0
2,4-D	8.393	8.294	8.494	733.030	705.000	4.0
2,4-DCAA	7.261	7.162	7.362	777.510	750.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028272.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.922	9.823	10.023	740.260	712.500	3.9
2,4-D	9.014	8.915	9.115	721.950	705.000	2.4
2,4-DCAA	7.764	7.665	7.865	776.030	750.000	3.5